



TOTAL SOLIDS - SM2540B

SUPERVISOR: apatelANALYST: AparnaDate: 08/21/2018OVEN TEMP1: 103 °CTEMP: 105 °CDATE & TIME IN: 08/20/2018 16:00 OUT: 08/21/2018 07:10OVEN TEMP2: 105 °CTEMP: 105 °CDATE & TIME IN: 08/21/2018 08:15 OUT: 08/21/2018 09:40Run Number: LB97416BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB97416BL	LB97416BL	83.4521	83.4521	100	83.4521	83.4521	83.4521	0.0000	0
2	J4555-07	TWP-02	83.4765	83.4765	100	86.3401	86.3401	86.3401	2.8636	28636
3	J4555-07DUP	TWP-02DUP	81.4903	81.4903	100	84.4980	84.4980	84.4980	3.0077	30077

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS082018

WorkList ID : 115839

Date : 8/20/2018 2:09:02 PM

11974116

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/24/2018	Water	J4555-07	TS	Cool 4 deg C	LIRO01	P63	TWP-02	08/17/2018	SM2540 B

Date/Time

Received by:

Relinquished by:

08/20/18, 15:40
AS
CP

Date/Time

Received by:

Relinquished by:

08/20/18, 16:20
CP
AS