

LB 97419


245.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

36806



A= 0.0000e+000

B= 2.7325e-004

C= 4.7456e-003

Rho= 0.9999393

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.018	0.018	48	0.000	48				
0.05	0.050	0.063	0.013	215	0.0 %	215				
0.2	0.200	0.228	0.028	817	0.0 %	817				
2.5	2.500	2.482	-0.018	9066	0.0 %	9066				
5.0	5.000	4.927	-0.073	18014	0.0 %	18014				
7.5	7.500	7.469	-0.031	27318	0.0 %	27318				
10.0	10.000	10.062	0.062	36806	0.0 %	36806				

LB97419 INSTRUMENT ID : CV1

Method: 245.1		Operator: Admin		Date of Analysis: 20 Aug 2018 16:31:39					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	48	-	0.0000	245.1	PPB	20 Aug 2018 16:34:05	S	Std
0.05 - 1	50.05	215	-	0.0500	245.1	PPB	20 Aug 2018 16:36:14	S	Std
0.2 - 1	50.2	817	-	0.2000	245.1	PPB	20 Aug 2018 16:38:22	S	Std
2.5 - 1	52.5	9066	-	2.5000	245.1	PPB	20 Aug 2018 16:40:30	S	Std
5.0 - 1	55	18014	-	5.0000	245.1	PPB	20 Aug 2018 16:42:38	S	Std
7.5 - 1	57.5	27318	-	7.5000	245.1	PPB	20 Aug 2018 16:44:45	S	Std
10.0 - 1	510	36806	-	10.0000	245.1	PPB	20 Aug 2018 16:52:01	S	Std
ICV18 - 1	ICV18	14527	3.9743	-245.1		PPB	20 Aug 2018 16:54:36	U	SMPL
ICB18 - 1	ICB18	-226	-0.0570	-245.1		PPB	20 Aug 2018 16:56:44	U	SMPL
CCV33 - 1	CCV33	18529	5.0678	-245.1		PPB	20 Aug 2018 16:58:52	U	SMPL
CCB33 - 1	CCB33	-115	-0.0267	-245.1		PPB	20 Aug 2018 17:00:58	U	SMPL
CRA - 1	CRA	213	0.0629	-245.1		PPB	20 Aug 2018 17:12:20	U	SMPL
HighStd - 1	HighStd	35855	9.8022	-245.1		PPB	20 Aug 2018 17:14:26	U	SMPL
ChkStd - 1	ChkStd	25414	6.9492	-245.1		PPB	20 Aug 2018 17:16:33	U	SMPL
PB112180BL - 1	PBW	-229	-0.0578	-245.1		PPB	20 Aug 2018 17:18:42	U	SMPL
PB112180BS - 1	LCSW	13419	3.6715	-245.1		PPB	20 Aug 2018 17:20:52	U	SMPL
J4525-01 - 1	081518-DBRI-E-D-B	-154	-0.0373	-245.1		PPB	20 Aug 2018 17:23:00	U	SMPL
J4525-02 - 1	081518-DBRI-E-D-S	-43	-0.0070	-245.1		PPB	20 Aug 2018 17:25:08	U	SMPL
J4525-03 - 1	081518-DBRI-E-U-B	-32	-0.0040	-245.1		PPB	20 Aug 2018 17:27:16	U	SMPL
J4525-04 - 1	081518-DBRI-E-U-S	-12	0.0015	-245.1		PPB	20 Aug 2018 17:29:25	U	SMPL
J4525-05 - 1	081518-TIPC-F-D-B	-23	-0.0015	-245.1		PPB	20 Aug 2018 17:31:31	U	SMPL
CCV34 - 1	CCV34	18024	4.9298	-245.1		PPB	20 Aug 2018 17:33:37	U	SMPL
CCB34 - 1	CCB34	-193	-0.0480	-245.1		PPB	20 Aug 2018 17:35:43	U	SMPL
J4525-06 - 1	081518-TIPC-F-D-S	-13	0.0012	-245.1		PPB	20 Aug 2018 17:37:51	U	SMPL
J4525-07 - 1	081518-TIPC-F-U-B	-12	0.0015	-245.1		PPB	20 Aug 2018 17:39:58	U	SMPL
J4525-08 - 1	081518-TIPC-F-U-S	-6	0.0031	-245.1		PPB	20 Aug 2018 17:42:04	U	SMPL
J4525-09 - 1	081518-SIDI-F-D-B	25	0.0116	-245.1		PPB	20 Aug 2018 17:44:11	U	SMPL
J4525-10 - 1	081518-SIDI-F-D-M	-1	0.0045	-245.1		PPB	20 Aug 2018 17:46:19	U	SMPL
J4525-11 - 1	081518-SIDI-F-D-S	5	0.0061	-245.1		PPB	20 Aug 2018 17:48:26	U	SMPL
J4525-12 - 1	081518-SIDI-F-U-B	34	0.0140	-245.1		PPB	20 Aug 2018 17:50:34	U	SMPL
J4525-13 - 1	081518-SIDI-F-U-M	-2	0.0042	-245.1		PPB	20 Aug 2018 17:52:42	U	SMPL
J4525-14 - 1	081518-SIDI-F-U-S	-2	0.0042	-245.1		PPB	20 Aug 2018 17:54:51	U	SMPL
J4525-14DUP - 1	081518-SIDI-F-U-SDUP	14	0.0086	-245.1		PPB	20 Aug 2018 17:56:58	U	SMPL
CCV35 - 1	CCV35	17857	4.8842	-245.1		PPB	20 Aug 2018 17:59:05	U	SMPL
CCB35 - 1	CCB35	-162	-0.0395	-245.1		PPB	20 Aug 2018 18:01:12	U	SMPL
J4525-14MS - 1	081518-SIDI-F-U-SMS	15263	4.1754	-245.1		PPB	20 Aug 2018 18:03:21	U	SMPL
J4525-14MSD - 1	081518-SIDI-F-U-SMSD	14612	3.9975	-245.1		PPB	20 Aug 2018 18:05:28	U	SMPL
J4525-14LX5 - 1		-107	-0.0245	-245.1		PPB	20 Aug 2018 18:07:35	U	SMPL
J4525-14A - 1		16556	4.5287	-245.1		PPB	20 Aug 2018 18:09:42	U	SMPL
CCV36 - 1	CCV36	18232	4.9867	-245.1		PPB	20 Aug 2018 18:11:50	U	SMPL
CCB36 - 1	CCB36	-186	-0.0461	-245.1		PPB	20 Aug 2018 18:13:58	U	SMPL