



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97421

Slope : 98.3

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/20/2018	14:10
2	CAL2	1	Water			21.7	7.01	08/20/2018	14:12
3	CAL3	1	Water			22.0	10.04	08/20/2018	14:14
4	ICV	1	Water			21.8	7.00	08/20/2018	14:16
5	CCV1	1	Water			21.9	2.00	08/20/2018	14:18
6	J4566-01	1	Water			22.1	7.06	08/20/2018	14:20
7	J4566-01DUP	1	Water			22.2	7.07	08/20/2018	14:22
8	CCV2	1	Water			22.0	12.00	08/20/2018	14:24

LB97421

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH082018

WorkList ID : 115837

Date : 8/20/2018 1:51:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2018	Water	J4566-01	pH	Cool 4 deg C	PSEG01	R22	NUTLEY-DRUM	08/20/2018	9040C

Date/Time 08/20/18 2:00
Received by: SP
Relinquished by: CP

Date/Time 08/20/18 2:30
Received by: CP
Relinquished by: SP