



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97422

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	08/20/2018	16:50
2	CAL2	1	Water			21.8	7.01	08/20/2018	16:52
3	CAL3	1	Water			21.9	10.04	08/20/2018	16:54
4	ICV	1	Water			22.1	7.00	08/20/2018	16:56
5	CCV1	1	Water			21.9	2.00	08/20/2018	16:58
6	J4569-01	1	Solid	20.02	20	23.6	6.03	08/20/2018	17:00
7	J4569-01DUP	1	Solid	20.03	20	23.7	6.04	08/20/2018	17:02
8	CCV2	1	Water			21.7	12.00	08/20/2018	17:04

WORKLIST(Hardcopy Internal Chain)

LB 97422

WorkList Name : pH082018-1

WorkList ID : 115846

Date : 8/20/2018 4:06:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2018	Solid	J4569-01	pH	Cool 4 deg C	TULL02	E22	FK-PS-01-037	08/20/2018	9045D

Date/Time 08/20/18 4:45
 Received by: SP
 Relinquished by: CP

Date/Time 08/20/18 5:10
 Received by: CP
 Relinquished by: SP