


7471B

LIB 97441
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44958



A= 0.0000e+000

B= 2.2414e-004

C= -2.2924e-002

Rho= 0.9997758

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.001	-0.001	96	0.000	96				
0.2	0.200	0.235	0.035	1151	0.0 %	1151				
2.5	2.500	2.431	-0.069	10949	0.0 %	10949				
5.0	5.000	5.108	0.108	22892	0.0 %	22892				
7.5	7.500	7.373	-0.127	32998	0.0 %	32998				
10.0	10.000	10.054	0.054	44958	0.0 %	44958				

LB97441 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Aug 2018 14:44:00

Sample ID	Extended ID	μ Abs.	Conc.	Stnd CondMethod	Units	Date	Type	Type
0.0 - 1	50	96	-	0.00007471B	PPB	21 Aug 2018 14:47:56	S	Std
0.2 - 1	50.2	1151	-	0.20007471B	PPB	21 Aug 2018 14:50:14	S	Std
2.5 - 1	52.5	10949	-	2.50007471B	PPB	21 Aug 2018 14:52:32	S	Std
5.0 - 1	55	22892	-	5.00007471B	PPB	21 Aug 2018 14:54:49	S	Std
7.5 - 1	57.5	32998	-	7.50007471B	PPB	21 Aug 2018 14:57:06	S	Std
10.0 - 1	510	44958	-	10.00007471B	PPB	21 Aug 2018 15:02:01	S	Std
ICV21 - 1	ICV21	18796	4.1900	-7471B	PPB	21 Aug 2018 15:04:43	U	SMPL
ICB21 - 1	ICB21	-114	-0.0485	-7471B	PPB	21 Aug 2018 15:06:59	U	SMPL
CCV46 - 1	CCV46	23254	5.1892	-7471B	PPB	21 Aug 2018 15:09:17	U	SMPL
CCB46 - 1	CCB46	-113	-0.0483	-7471B	PPB	21 Aug 2018 15:11:32	U	SMPL
CRA - 1	CRA	1116	0.2272	-7471B	PPB	21 Aug 2018 15:16:21	U	SMPL
HighStd - 1	HighStd	45058	10.0763	-7471B	PPB	21 Aug 2018 15:18:37	U	SMPL
ChkStd - 1	ChkStd	32423	7.2443	-7471B	PPB	21 Aug 2018 15:20:53	U	SMPL
PB112230BL - 1	PBS	-186	-0.0646	-7471B	PPB	21 Aug 2018 15:23:11	U	SMPL
PB112230BS - 1	LCSS	16096	3.5848	-7471B	PPB	21 Aug 2018 15:25:29	U	SMPL
J4271-09 - 1	HR-06-082018-A	12626	2.8071	-7471B	PPB	21 Aug 2018 15:27:47	U	SMPL
J4271-09DUP - 1	HR-06-082018-ADUP	11243	2.4971	-7471B	PPB	21 Aug 2018 15:30:04	U	SMPL
J4271-09MS - 1	HR-06-082018-AMS	29734	6.6416	-7471B	PPB	21 Aug 2018 15:32:23	U	SMPL
J4271-09MSD - 1	HR-06-082018-AMSD	29606	6.6129	-7471B	PPB	21 Aug 2018 15:34:42	U	SMPL
J4271-11 - 1	HR-06-082018-B	10511	2.3330	-7471B	PPB	21 Aug 2018 15:37:00	U	SMPL
CCV47 - 1	CCV47	22990	5.1300	-7471B	PPB	21 Aug 2018 15:39:18	U	SMPL
CCB47 - 1	CCB47	-142	-0.0548	-7471B	PPB	21 Aug 2018 15:41:35	U	SMPL
J4271-13 - 1	HR-06-082018-C	4778	1.0480	-7471B	PPB	21 Aug 2018 15:43:52	U	SMPL
J4275-09 - 1	CL-01-082018-A	4748	1.0413	-7471B	PPB	21 Aug 2018 15:46:08	U	SMPL
J4275-11 - 1	CL-01-082018-B	1757	0.3709	-7471B	PPB	21 Aug 2018 15:48:24	U	SMPL
J4275-13 - 1	CL-01-082018-C	3585	0.7806	-7471B	PPB	21 Aug 2018 15:50:41	U	SMPL
J4275-15 - 1	CL-01-082018-D	3509	0.7636	-7471B	PPB	21 Aug 2018 15:52:57	U	SMPL
J4275-17 - 1	CL-01-082018-E	2442	0.5244	-7471B	PPB	21 Aug 2018 15:55:14	U	SMPL
J4553-01 - 1	EPE-P-9(20-22)	734	0.1416	-7471B	PPB	21 Aug 2018 15:57:31	U	SMPL
J4553-02 - 1	EPE-P-9(30-36)	108	0.0013	-7471B	PPB	21 Aug 2018 15:59:49	U	SMPL
J4555-03 - 1	SB-01-COMP	6189	1.3643	-7471B	PPB	21 Aug 2018 16:02:06	U	SMPL
J4555-05 - 1	SB-02-COMP	14227	3.1659	-7471B	PPB	21 Aug 2018 16:04:23	U	SMPL
CCV48 - 1	CCV48	23024	5.1376	-7471B	PPB	21 Aug 2018 16:06:39	U	SMPL
CCB48 - 1	CCB48	-136	-0.0534	-7471B	PPB	21 Aug 2018 16:08:55	U	SMPL
J4556-02 - 1	20180530-COMPOSITE-(ABC)	41123	9.1943	-7471B	PPB	21 Aug 2018 16:11:12	U	SMPL
J4556-03 - 1	20180531-COMPOSITE-(DEF)	23151	5.1661	-7471B	PPB	21 Aug 2018 16:13:29	U	SMPL
J4556-04 - 1	20180532-COMPOSITE-(GHI)	35184	7.8632	-7471B	PPB	21 Aug 2018 16:15:46	U	SMPL
J4556-08 - 1	20180533-AMENDED-COMPOSITE-(ABC)	32523	7.2667	-7471B	PPB	21 Aug 2018 16:18:04	U	SMPL
J4556-09 - 1	20180534-AMENDED-COMPOSITE-(DEF)	18466	4.1160	-7471B	PPB	21 Aug 2018 16:20:23	U	SMPL
J4556-10 - 1	20180535-AMENDED-COMPOSITE-(GHI)	30297	6.7678	-7471B	PPB	21 Aug 2018 16:22:42	U	SMPL
PB112231BL - 1	PBS	-25	-0.0285	-7471B	PPB	21 Aug 2018 16:24:59	U	SMPL
PB112231BS - 1	LCSS	15561	3.4649	-7471B	PPB	21 Aug 2018 16:27:18	U	SMPL
J4559-02 - 1	RA-081718-FL-046	3434	0.7468	-7471B	PPB	21 Aug 2018 16:29:36	U	SMPL
J4559-04 - 1	RA-081718-SSW-75-048	4188	0.9158	-7471B	PPB	21 Aug 2018 16:31:52	U	SMPL
CCV49 - 1	CCV49	22683	5.0612	-7471B	PPB	21 Aug 2018 16:34:08	U	SMPL
CCB49 - 1	CCB49	-94	-0.0440	-7471B	PPB	21 Aug 2018 16:36:25	U	SMPL
J4559-05 - 1	RA-081718-WSW-20-049	7803	1.7260	-7471B	PPB	21 Aug 2018 16:38:42	U	SMPL
J4559-06 - 1	RA-081718-SSW-115-050	17951	4.0006	-7471B	PPB	21 Aug 2018 16:40:59	U	SMPL
J4559-07 - 1	RA-081718-SSW-125-051	5993	1.3203	-7471B	PPB	21 Aug 2018 16:43:15	U	SMPL
J4559-08 - 1	RA-081718-SSW-145-052	4050	0.8848	-7471B	PPB	21 Aug 2018 16:45:31	U	SMPL
J4559-09 - 1	RA-081718-WSW-20-053	10676	2.3700	-7471B	PPB	21 Aug 2018 16:47:47	U	SMPL
J4559-09DUP - 1	RA-081718-WSW-20-053DUP	10112	2.2436	-7471B	PPB	21 Aug 2018 16:50:03	U	SMPL
J4559-10 - 1	J4559-09MS	28586	6.3843	-7471B	PPB	21 Aug 2018 16:52:19	U	SMPL
J4559-11 - 1	J4559-09MSD	26039	5.8134	-7471B	PPB	21 Aug 2018 16:54:35	U	SMPL
J4561-02 - 1	HUDSON-RIVER-PARK-TRUST-WC-1-COMPOSITE	67722	15.1562	-7471B	PPB	21 Aug 2018 16:56:51	U	SMPL
J4569-01 - 1	FK-PS-01-037	299	0.0441	-7471B	PPB	21 Aug 2018 17:01:57	U	SMPL
CCV50 - 1	CCV50	22937	5.1181	-7471B	PPB	21 Aug 2018 17:04:14	U	SMPL
CCB50 - 1	CCB50	-86	-0.0422	-7471B	PPB	21 Aug 2018 17:06:32	U	SMPL
J4572-01 - 1	F-1	1575	0.3301	-7471B	PPB	21 Aug 2018 17:08:49	U	SMPL
J4573-03 - 1	T25-COMP	63657	14.2451	-7471B	PPB	21 Aug 2018 17:11:05	U	SMPL
J4573-07 - 1	T26-COMP	70471	15.7724	-7471B	PPB	21 Aug 2018 17:13:21	U	SMPL
J4573-11 - 1	T27-COMP	42985	9.6117	-7471B	PPB	21 Aug 2018 17:15:38	U	SMPL
J4271-09LX5 - 1		2100	0.4478	-7471B	PPB	21 Aug 2018 17:17:57	U	SMPL
J4271-09A - 1		28679	6.4052	-7471B	PPB	21 Aug 2018 17:20:16	U	SMPL
J4559-09LX5 - 1		2280	0.4881	-7471B	PPB	21 Aug 2018 17:22:34	U	SMPL
J4559-09A - 1		29989	6.6988	-7471B	PPB	21 Aug 2018 17:24:51	U	SMPL
CCV51 - 1	CCV51	22891	5.1078	-7471B	PPB	21 Aug 2018 17:27:10	U	SMPL
CCB51 - 1	CCB51	-108	-0.0471	-7471B	PPB	21 Aug 2018 17:29:28	U	SMPL
J4561-02DLX2 - 1		34760	7.7681	-7471B	PPB	21 Aug 2018 17:31:45	U	SMPL
J4573-03DLX2 - 1		32151	7.1834	-7471B	PPB	21 Aug 2018 17:34:02	U	SMPL
J4573-07DLX2 - 1		36222	8.0958	-7471B	PPB	21 Aug 2018 17:36:19	U	SMPL
CCV52 - 1	CCV52	23433	5.2293	-7471B	PPB	21 Aug 2018 17:38:37	U	SMPL
CCB52 - 1	CCB52	-93	-0.0438	-7471B	PPB	21 Aug 2018 17:40:54	U	SMPL