



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97443

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	08/21/2018	09:40
2	J4555-07		1	8	>100.0	>212.0	>212.0	08/21/2018	10:15
3	J4566-02		1	6	>100.0	>212.0	>212.0	08/21/2018	10:46
4	J4566-02DUP		1	6	>100.0	>212.0	>212.0	08/21/2018	11:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 97 443

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp82118

WorkList ID : 115876

Date : 8/21/2018 8:59:25 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/24/2018	Water	J4555-07	Flash Point	Cool 4 deg C	LIRO01	P63	TWP-02	08/17/2018	1010A
08/23/2018	Water	J4566-02	Flash Point	Cool 4 deg C	PSEG01	R22	NUTLEY-DRUM	08/20/2018	1010A

Date/Time 08/21/18 9:10 AM
Received by: KS
Relinquished by: CP

Date/Time 08/21/18 11:25 AM
Received by: CP
Relinquished by: KS