

21397447

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/21/2018 16:08

Test: Phenolics-mg/L

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.994	0.0	0.109	
ICB1	0.015	0.0	0.005	
CCV1	0.971	0.0	0.106	
CCB1	0.010	0.0	0.004	
PB112223BL	0.007	0.0	0.004	
PB112223BS	1.001	0.0	0.109	
J4555-07	0.020	0.0	0.006	
J4555-07DUP	0.015	0.0	0.005	
J4555-07MS	0.895	0.0	0.098	
J4555-07MSD	0.902	0.0	0.099	
CCV2	0.965	0.0	0.106	
CCB2	0.004	0.0	0.004	

N 12
Mean 0.483
SD 0.4934
CV% 102.12

Aquakem v. 7.2AQ1

Results from time period:

Tue Aug 21 15:45:32 2018

Tue Aug 21 16:07:26 2018

✓

Sample Id	Sam/Ctr/c/	Test short name	Test type	Result	Result unit	Result date and time	Stat
0.05PPM	A	Phenolics-	P	0.0522	mg/l	8/21/2018 15:45:32	
0.1PPM	A	Phenolics-	P	0.1059	mg/l	8/21/2018 15:45:33	
0.25PPM	A	Phenolics-	P	0.2508	mg/l	8/21/2018 15:45:34	
0.50PPM	A	Phenolics-	P	0.5005	mg/l	8/21/2018 15:45:35	
1.0PPM	A	Phenolics-	P	0.9824	mg/l	8/21/2018 15:45:36	
2.0PPM	A	Phenolics-	P	2.0082	mg/l	8/21/2018 15:45:37	
ICV1	S	Phenolics-	P	0.9937	mg/l	8/21/2018 16:07:15	
ICB1	S	Phenolics-	P	0.0147	mg/l	8/21/2018 16:07:16	
CCV1	S	Phenolics-	P	0.9715	mg/l	8/21/2018 16:07:17	
CCB1	S	Phenolics-	P	0.0097	mg/l	8/21/2018 16:07:18	
PB112223BL	S	Phenolics-	P	0.0067	mg/l	8/21/2018 16:07:19	
PB112223BS	S	Phenolics-	P	1.001	mg/l	8/21/2018 16:07:20	
J4555-07	S	Phenolics-	P	0.02	mg/l	8/21/2018 16:07:21	
J4555-07DUP	S	Phenolics-	P	0.0152	mg/l	8/21/2018 16:07:22	
J4555-07MS	S	Phenolics-	P	0.8951	mg/l	8/21/2018 16:07:23	
J4555-07MSD	S	Phenolics-	P	0.902	mg/l	8/21/2018 16:07:24	
CCV2	S	Phenolics-	P	0.9651	mg/l	8/21/2018 16:07:25	
CCB2	S	Phenolics-	P	0.0041	mg/l	8/21/2018 16:07:26	

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : AS

8/21/2018 15:45


Test Phenolics-

Accepted 8/21/2018 15:45

Factor 9.444
 Bias 0.003

Coeff. of det. 0.999851

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	Phenol-2	0.009	0.0522	0.0500	
2	Phenol-2	0.015	0.1059	0.1000	
3	Phenol-2	0.030	0.2508	0.2500	
4	Phenol-2	0.056	0.5005	0.5000	
5	Phenol-2	0.107	0.9824	1.0000	
6	Phenol-2	0.216	2.0082	2.0000	