

LB 97460


7470A

INSTRUMENT ID: CXL

Linear

 μ Abs.:

36005



A= 0.0000e+000

B= 2.7675e-004

C= -7.9973e-002

Rho= 0.9997263

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.035	-0.035	161	0.000	161				
0.2	0.200	0.240	0.040	1157	0.0 %	1157				
2.5	2.500	2.420	-0.080	9033	0.0 %	9033				
5.0	5.000	5.051	0.051	18541	0.0 %	18541				
7.5	7.500	7.639	0.139	27893	0.0 %	27893				
10.0	10.000	9.884	-0.116	36005	0.0 %	36005				

LB97460 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Aug 2018 09:20:51

Sample ID	Extended ID	μ Abs	Conc	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50.0	161	-	0.0007470A	PPB	22 Aug 2018 09:26:51	S	Std
0.2 - 1	50.2	1157	-	0.2007470A	PPB	22 Aug 2018 09:28:59	S	Std
2.5 - 1	52.5	9033	-	2.5007470A	PPB	22 Aug 2018 09:31:08	S	Std
5.0 - 1	55	18541	-	5.0007470A	PPB	22 Aug 2018 09:33:16	S	Std
7.5 - 1	57.5	27893	-	7.5007470A	PPB	22 Aug 2018 09:35:23	S	Std
10.0 - 1	59	36005	-	10.0007470A	PPB	22 Aug 2018 09:37:31	S	Std
ICV22 - 1	ICV22	14387	3.9017	-7470A	PPB	22 Aug 2018 09:40:50	U	SMPL
ICB22 - 1	ICB22	79	-0.0581	-7470A	PPB	22 Aug 2018 09:42:57	U	SMPL
CCV53 - 1	CCV53	18222	4.9630	-7470A	PPB	22 Aug 2018 09:45:04	U	SMPL
CCB53 - 1	CCB53	61	-0.0631	-7470A	PPB	22 Aug 2018 09:47:11	U	SMPL
CRA - 1	CRA	1124	0.2311	-7470A	PPB	22 Aug 2018 09:49:17	U	SMPL
HighStd - 1	HighStd	35895	9.8540	-7470A	PPB	22 Aug 2018 09:51:24	U	SMPL
ChkStd - 1	ChkStd	25309	6.9243	-7470A	PPB	22 Aug 2018 09:53:30	U	SMPL
PB112270BL - 1	PBW	166	-0.0340	-7470A	PPB	22 Aug 2018 10:00:12	U	SMPL
PB112270BS - 1	LCSW	13435	3.6382	-7470A	PPB	22 Aug 2018 10:03:28	U	SMPL
PB112270TB - 1	PB112226TB	46	-0.0672	-7470A	PPB	22 Aug 2018 10:05:36	U	SMPL
J4271-10 - 1	HR-06-082018-A	139	-0.0415	-7470A	PPB	22 Aug 2018 10:07:44	U	SMPL
J4271-12 - 1	HR-06-082018-B	142	-0.0407	-7470A	PPB	22 Aug 2018 10:09:52	U	SMPL
J4271-14 - 1	HR-06-082018-C	149	-0.0387	-7470A	PPB	22 Aug 2018 10:12:01	U	SMPL
J4271-14DUP - 1	HR-06-082018-CDUP	135	-0.0426	-7470A	PPB	22 Aug 2018 10:14:08	U	SMPL
CCV54 - 1	CCV54	18093	4.9273	-7470A	PPB	22 Aug 2018 10:16:14	U	SMPL
CCB54 - 1	CCB54	10	-0.0772	-7470A	PPB	22 Aug 2018 10:18:21	U	SMPL
J4271-14MS - 1	HR-06-082018-CMS	12953	3.5048	-7470A	PPB	22 Aug 2018 10:22:57	U	SMPL
J4271-14MSD - 1	HR-06-082018-CMSD	13034	3.5272	-7470A	PPB	22 Aug 2018 10:25:03	U	SMPL
J4275-10 - 1	CL-01-082018-A	30	-0.0717	-7470A	PPB	22 Aug 2018 10:27:09	U	SMPL
J4275-12 - 1	CL-01-082018-B	117	-0.0476	-7470A	PPB	22 Aug 2018 10:29:18	U	SMPL
J4275-14 - 1	CL-01-082018-C	149	-0.0387	-7470A	PPB	22 Aug 2018 10:31:25	U	SMPL
J4275-16 - 1	CL-01-082018-D	179	-0.0304	-7470A	PPB	22 Aug 2018 10:33:33	U	SMPL
J4275-18 - 1	CL-01-082018-E	161	-0.0354	-7470A	PPB	22 Aug 2018 10:35:41	U	SMPL
J4555-04 - 1	SB-01-COMP	205	-0.0232	-7470A	PPB	22 Aug 2018 10:37:49	U	SMPL
J4555-06 - 1	SB-02-COMP	205	-0.0232	-7470A	PPB	22 Aug 2018 10:39:56	U	SMPL
J4560-02 - 1	CB-0818-S-TCLP-GRID61-2	173	-0.0321	-7470A	PPB	22 Aug 2018 10:42:05	U	SMPL
CCV55 - 1	CCV55	18203	4.9577	-7470A	PPB	22 Aug 2018 10:44:13	U	SMPL
CCB55 - 1	CCB55	36	-0.0700	-7470A	PPB	22 Aug 2018 10:46:20	U	SMPL
J4561-04 - 1	HUDSON-RIVER-PARK-TRUST-WC-1-COMPOSITE	205	-0.0232	-7470A	PPB	22 Aug 2018 10:48:28	U	SMPL
J4573-04 - 1	T25-COMP	711	0.1168	-7470A	PPB	22 Aug 2018 10:50:35	U	SMPL
J4573-08 - 1	T26-COMP	212	-0.0213	-7470A	PPB	22 Aug 2018 10:52:42	U	SMPL
J4573-12 - 1	T27-COMP	227	-0.0171	-7470A	PPB	22 Aug 2018 10:54:49	U	SMPL
PB112271BL - 1	PBW	140	-0.0412	-7470A	PPB	22 Aug 2018 10:56:56	U	SMPL
PB112271BS - 1	LCSW	13041	3.5292	-7470A	PPB	22 Aug 2018 10:59:04	U	SMPL
PB112271TB - 1	PB112228TB	89	-0.0553	-7470A	PPB	22 Aug 2018 11:01:12	U	SMPL
J4566-02 - 1	NUTLEY-DRUM	632	0.0949	-7470A	PPB	22 Aug 2018 11:06:45	U	SMPL
J4566-02DUP - 1	NUTLEY-DRUMDUP	629	0.0941	-7470A	PPB	22 Aug 2018 11:10:53	U	SMPL
J4566-02MS - 1	NUTLEY-DRUMMS	13017	3.5225	-7470A	PPB	22 Aug 2018 11:13:01	U	SMPL
CCV56 - 1	CCV56	18800	5.1230	-7470A	PPB	22 Aug 2018 11:15:08	U	SMPL
CCB56 - 1	CCB56	20	-0.0744	-7470A	PPB	22 Aug 2018 11:17:16	U	SMPL
J4566-02MSD - 1	NUTLEY-DRUMMSD	13640	3.6949	-7470A	PPB	22 Aug 2018 11:19:25	U	SMPL
J4271-14LX5 - 1		43	-0.0681	-7470A	PPB	22 Aug 2018 11:21:32	U	SMPL
J4271-14A - 1		14107	3.8242	-7470A	PPB	22 Aug 2018 11:23:40	U	SMPL
J4566-02LX5 - 1		225	-0.0177	-7470A	PPB	22 Aug 2018 11:25:47	U	SMPL
J4566-02A - 1		15127	4.1065	-7470A	PPB	22 Aug 2018 11:27:55	U	SMPL
CCV57 - 1	CCV57	18675	5.0884	-7470A	PPB	22 Aug 2018 11:30:03	U	SMPL
CCB57 - 1	CCB57	9	-0.0775	-7470A	PPB	22 Aug 2018 11:32:12	U	SMPL