



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97462

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/21/2018	14:15
2	CAL2	1	Water			22.0	7.01	08/21/2018	14:17
3	CAL3	1	Water			21.8	10.04	08/21/2018	14:19
4	ICV	1	Water			21.8	7.00	08/21/2018	14:21
5	CCV1	1	Water			22.0	2.00	08/21/2018	14:23
6	J4587-01	1	Water			23.9	5.29	08/21/2018	14:25
7	J4587-01DUP	1	Water			24.0	5.30	08/21/2018	14:27
8	CCV2	1	Water			21.8	12.00	08/21/2018	14:29

WORKLIST(Hardcopy Internal Chain)

UB 97462

WorkList Name : pH082118

WorkList ID : 115898

Date : 8/21/2018 2:09:19 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/20/2018	Water	J4587-01 D	pH	Cool 4 deg C	STER01	E21	POOL-WATER	08/20/2018	SM 4500-PH B

Date/Time 08/21/18 2:12
 Received by: SP
 Relinquished by: CP

Date/Time 08/21/18 2:45
 Received by: CP
 Relinquished by: SP