

245.1


L13 97468

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34495



A= 0.0000e+000

B= 2.9215e-004

C= -6.0638e-003

Rho= 0.9998964

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	4	0.000	4				
0.05	0.050	0.058	0.008	220	0.0 %	220				
0.2	0.200	0.246	0.046	863	0.0 %	863				
2.5	2.500	2.456	-0.044	8427	0.0 %	8427				
5.0	5.000	5.024	0.024	17216	0.0 %	17216				
7.5	7.500	7.399	-0.101	25348	0.0 %	25348				
10.0	10.000	10.072	0.072	34495	0.0 %	34495				

LB97468 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 22 Aug 2018 13:34:18

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	4	-	0.0000245.1	PPB	22 Aug 2018 13:37:53	S	Std
0.05 - 1	50.05	220	-	0.0500245.1	PPB	22 Aug 2018 13:40:01	S	Std
0.2 - 1	50.2	863	-	0.2000245.1	PPB	22 Aug 2018 13:42:10	S	Std
2.5 - 1	52.5	8427	-	2.5000245.1	PPB	22 Aug 2018 13:44:18	S	Std
5.0 - 1	55	17216	-	5.0000245.1	PPB	22 Aug 2018 13:46:26	S	Std
7.5 - 1	57.5	25348	-	7.5000245.1	PPB	22 Aug 2018 13:48:33	S	Std
10.0 - 1	510	34495	-	10.0000245.1	PPB	22 Aug 2018 13:50:42	S	Std
ICV23 - 1	ICV23	13686	3.9923	-245.1	PPB	22 Aug 2018 13:53:34	U	SMPL
ICB23 - 1	ICB23	-165	-0.0543	-245.1	PPB	22 Aug 2018 13:55:40	U	SMPL
CCV58 - 1	CCV58	17326	5.0558	-245.1	PPB	22 Aug 2018 13:57:48	U	SMPL
CCB58 - 1	CCB58	-177	-0.0578	-245.1	PPB	22 Aug 2018 13:59:55	U	SMPL
CRA - 1	CRA	185	0.0480	-245.1	PPB	22 Aug 2018 14:02:03	U	SMPL
HighStd - 1	HighStd	34492	10.0708	-245.1	PPB	22 Aug 2018 14:04:09	U	SMPL
ChkStd - 1	ChkStd	23285	6.7967	-245.1	PPB	22 Aug 2018 14:06:16	U	SMPL
PB112272BL - 1	PBW	-72	-0.0271	-245.1	PPB	22 Aug 2018 14:08:26	U	SMPL
PB112272BS - 1	LCSW	12880	3.7568	-245.1	PPB	22 Aug 2018 14:10:34	U	SMPL
J4540-05 - 1	081618-DBRI-E-D-B	81	0.0176	-245.1	PPB	22 Aug 2018 14:12:42	U	SMPL
J4540-06 - 1	081618-DBRI-E-D-S	170	0.0436	-245.1	PPB	22 Aug 2018 14:14:50	U	SMPL
J4540-07 - 1	081618-DBRI-E-U-B	244	0.0652	-245.1	PPB	22 Aug 2018 14:16:59	U	SMPL
J4540-08 - 1	081618-DBRI-E-U-S	192	0.0500	-245.1	PPB	22 Aug 2018 14:19:07	U	SMPL
J4568-01 - 1	081718-SIDI-E-D-B	224	0.0594	-245.1	PPB	22 Aug 2018 14:21:14	U	SMPL
CCV59 - 1	CCV59	17493	5.1045	-245.1	PPB	22 Aug 2018 14:23:21	U	SMPL
CCB59 - 1	CCB59	-154	-0.0511	-245.1	PPB	22 Aug 2018 14:25:27	U	SMPL
J4568-02 - 1	081718-SIDI-E-D-S	171	0.0439	-245.1	PPB	22 Aug 2018 14:27:35	U	SMPL
J4568-03 - 1	081718-SIDI-E-U-B	205	0.0538	-245.1	PPB	22 Aug 2018 14:29:42	U	SMPL
J4568-04 - 1	081718-SIDI-E-U-S	226	0.0600	-245.1	PPB	22 Aug 2018 14:31:49	U	SMPL
J4568-05 - 1	081718-DBRI-E-D-B	178	0.0459	-245.1	PPB	22 Aug 2018 14:33:56	U	SMPL
J4568-06 - 1	081718-DBRI-E-D-S	225	0.0597	-245.1	PPB	22 Aug 2018 14:36:04	U	SMPL
J4568-07 - 1	081718-DBRI-E-U-B	213	0.0562	-245.1	PPB	22 Aug 2018 14:38:12	U	SMPL
J4568-08 - 1	081718-DBRI-E-U-S	95	0.0217	-245.1	PPB	22 Aug 2018 14:40:19	U	SMPL
J4568-08DUP - 1	081718-DBRI-E-U-SDUP	91	0.0205	-245.1	PPB	22 Aug 2018 14:42:28	U	SMPL
J4568-08MS - 1	081718-DBRI-E-U-SMS	14584	4.2547	-245.1	PPB	22 Aug 2018 14:44:36	U	SMPL
J4568-08MSD - 1	081718-DBRI-E-U-SMSD	14401	4.2012	-245.1	PPB	22 Aug 2018 14:46:43	U	SMPL
CCV60 - 1	CCV60	17483	5.1016	-245.1	PPB	22 Aug 2018 14:48:50	U	SMPL
CCB60 - 1	CCB60	-171	-0.0560	-245.1	PPB	22 Aug 2018 14:50:57	U	SMPL
J4568-08LX5 - 1		-5	-0.0075	-245.1	PPB	22 Aug 2018 14:53:06	U	SMPL
J4568-08A - 1		14811	4.3210	-245.1	PPB	22 Aug 2018 14:55:13	U	SMPL
CCV61 - 1	CCV61	17431	5.0864	-245.1	PPB	22 Aug 2018 14:57:20	U	SMPL
CCB61 - 1	CCB61	-202	-0.0651	-245.1	PPB	22 Aug 2018 14:59:28	U	SMPL