

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/22/2018 14:08

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	100.998	0.0	0.096	
ICB1	1.525	0.0	0.005	
CCV1	253.701	0.0	0.236	
CCB1	1.359	0.0	0.005	
PB112286BL	0.782	0.0	0.004	
J4276-04	0.098	0.0	0.004	
J4276-04DUP	-0.352	0.0	0.003	
J4278-08	0.557	0.0	0.004	
J4278-10	0.185	0.0	0.004	
J4278-12	0.190	0.0	0.004	
J4283-04	0.079	0.0	0.004	
J4283-06	0.336	0.0	0.004	
J4580-04	0.595	0.0	0.004	
J4584-01	0.665	0.0	0.004	
CCV2	257.876	0.0	0.240	
CCB2	1.074	0.0	0.005	
J4585-02	0.029	0.0	0.004	
CCV3	259.484	0.0	0.241	
CCB3	1.184	0.0	0.005	

N	19
Mean	46.335
SD	96.4980
CV%	208.26

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/22/2018 11:24


Test Total CN

Accepted 8/22/2018 11:24

Factor 1092
Bias 0.004

Coeff. of det. 0.999996

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.004	-0.0443	0.0000	
2	5.0PPBCN	0.008	4.6234	5.0000	
3	10PPBCN	0.013	9.8368	10.0000	
4	50PPBCN	0.049	49.8274	50.0000	
5	100PPBCN	0.096	100.7676	100.0000	
6	250PPBCN	0.233	250.2366	250.0000	
7	500PPBCN	0.461	499.7524	500.0000	

Aquakem v. 7.2AQ1

AS

Results from time period:

Wed Aug 22 14:02:11 2018

Wed Aug 22 14:08:29 2018

Sample Id	Sam/Ctr/c/	Test short narr	Test type	Result	Result unit	Result date and time	Stat
0.OPPBCN	A	Reactive CN	P	-0.0443	µg/l	8/22/2018 11:22:55	
5.OPPBCN	A	Reactive CN	P	4.6234	µg/l	8/22/2018 11:22:56	
10PPBCN	A	Reactive CN	P	9.8368	µg/l	8/22/2018 11:22:57	
50PPBCN	A	Reactive CN	P	49.8274	µg/l	8/22/2018 11:22:58	
100PPBCN	A	Reactive CN	P	100.7676	µg/l	8/22/2018 11:22:59	
250PPBCN	A	Reactive CN	P	250.2366	µg/l	8/22/2018 11:23:00	
500PPBCN	A	Reactive CN	P	499.7524	µg/l	8/22/2018 11:23:01	
ICV1	S	Reactive CN	P	100.9976	µg/l	8/22/2018 14:02:11	
ICB1	S	Reactive CN	P	1.5246	µg/l	8/22/2018 14:02:12	
CCV1	S	Reactive CN	P	253.7009	µg/l	8/22/2018 14:02:13	
CCB1	S	Reactive CN	P	1.3592	µg/l	8/22/2018 14:02:14	
PB112286BL	S	Reactive CN	P	0.7821	µg/l	8/22/2018 14:02:15	
J4276-04	S	Reactive CN	P	0.0981	µg/l	8/22/2018 14:02:16	
J4276-04DUP	S	Reactive CN	P	-0.3524	µg/l	8/22/2018 14:02:17	
J4278-08	S	Reactive CN	P	0.5569	µg/l	8/22/2018 14:02:18	
J4278-10	S	Reactive CN	P	0.1847	µg/l	8/22/2018 14:02:19	
J4278-12	S	Reactive CN	P	0.1903	µg/l	8/22/2018 14:02:20	
J4283-04	S	Reactive CN	P	0.0791	µg/l	8/22/2018 14:02:21	
J4283-06	S	Reactive CN	P	0.3361	µg/l	8/22/2018 14:08:22	
J4580-04	S	Reactive CN	P	0.5953	µg/l	8/22/2018 14:08:23	
J4584-01	S	Reactive CN	P	0.6652	µg/l	8/22/2018 14:08:24	
CCV2	S	Reactive CN	P	257.8763	µg/l	8/22/2018 14:08:25	
CCB2	S	Reactive CN	P	1.0739	µg/l	8/22/2018 14:08:26	
J4585-02	S	Reactive CN	P	0.0293	µg/l	8/22/2018 14:08:27	
CCV3	S	Reactive CN	P	259.4836	µg/l	8/22/2018 14:08:28	
CCB3	S	Reactive CN	P	1.1838	µg/l	8/22/2018 14:08:29	