


7471B

LB 97473
INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.2307e-004

C= -8.8832e-003

Rho= 0.9999166

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	40	0.000	40				
0.2	0.200	0.195	-0.005	914	0.0 %	914				
2.5	2.500	2.493	-0.007	11218	0.0 %	11218				
5.0	5.000	5.067	0.067	22754	0.0 %	22754				
7.5	7.500	7.411	-0.089	33262	0.0 %	33262				
10.0	10.000	10.034	0.034	45021	0.0 %	45021				

LB97473 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 22 Aug 2018 15:17:29

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	40	-	0.0000	7471B	PPB	22 Aug 2018 16:01:20	S	Std
0.2 - 1	50.2	914	-	0.2000	7471B	PPB	22 Aug 2018 16:03:37	S	Std
2.5 - 1	52.5	11218	-	2.5000	7471B	PPB	22 Aug 2018 16:05:55	S	Std
5.0 - 1	55	22754	-	5.0000	7471B	PPB	22 Aug 2018 16:08:13	S	Std
7.5 - 1	57.5	33262	-	7.5000	7471B	PPB	22 Aug 2018 16:10:30	S	Std
10.0 - 1	510	45021	-	10.0000	7471B	PPB	22 Aug 2018 16:12:47	S	Std
ICV24 - 1	ICV24	17773	3.9557	-	7471B	PPB	22 Aug 2018 16:15:32	U	SMPL
ICB24 - 1	ICB24	127	0.0194	-	7471B	PPB	22 Aug 2018 16:17:48	U	SMPL
CCV62 - 1	CCV62	22984	5.1181	-	7471B	PPB	22 Aug 2018 16:20:04	U	SMPL
CCB62 - 1	CCB62	44	0.0009	-	7471B	PPB	22 Aug 2018 16:22:20	U	SMPL
CRA - 1	CRA	937	0.2001	-	7471B	PPB	22 Aug 2018 16:24:37	U	SMPL
HighStd - 1	HighStd	44946	10.0171	-	7471B	PPB	22 Aug 2018 16:26:52	U	SMPL
ChkStd - 1	ChkStd	31991	7.1273	-	7471B	PPB	22 Aug 2018 16:29:09	U	SMPL
PB112275BL - 1	PBS	-89	-0.0287	-	7471B	PPB	22 Aug 2018 16:31:27	U	SMPL
PB112275BS - 1	LCSS	15942	3.5473	-	7471B	PPB	22 Aug 2018 16:33:46	U	SMPL
J4276-03 - 1	OK-01-082118	3170	0.6982	-	7471B	PPB	22 Aug 2018 16:36:03	U	SMPL
J4278-07 - 1	EO-01-082118-A	1720	0.3748	-	7471B	PPB	22 Aug 2018 16:38:20	U	SMPL
J4278-09 - 1	EO-01-082118-B	2556	0.5613	-	7471B	PPB	22 Aug 2018 16:40:38	U	SMPL
J4278-11 - 1	EO-01-082118-C	1797	0.3920	-	7471B	PPB	22 Aug 2018 16:42:56	U	SMPL
J4283-03 - 1	TR-05-082118	11254	2.5015	-	7471B	PPB	22 Aug 2018 16:45:11	U	SMPL
CCV63 - 1	CCV63	22511	5.0126	-	7471B	PPB	22 Aug 2018 16:47:28	U	SMPL
CCB63 - 1	CCB63	42	0.0005	-	7471B	PPB	22 Aug 2018 16:49:43	U	SMPL
J4283-05 - 1	TR-04-082118	10855	2.4125	-	7471B	PPB	22 Aug 2018 16:52:00	U	SMPL
J4580-01 - 1	RN-SOIL	129164	28.8035	-	7471B	PPB	22 Aug 2018 16:54:15	U	SMPL
J4583-02 - 1	TOD1-SB-01-02-03-0-52	536	0.1107	-	7471B	PPB	22 Aug 2018 16:59:39	U	SMPL
J4583-02DUP - 1	TOD1-SB-01-02-03-0-52DUP	475	0.0971	-	7471B	PPB	22 Aug 2018 17:01:55	U	SMPL
J4583-03 - 1	J4583-02MS	19805	4.4090	-	7471B	PPB	22 Aug 2018 17:09:20	U	SMPL
J4583-04 - 1	J4583-02MSD	19569	4.3563	-	7471B	PPB	22 Aug 2018 17:11:36	U	SMPL
J4583-05 - 1	TOD1-SB-101-02-03-0-52	321	0.0627	-	7471B	PPB	22 Aug 2018 17:13:54	U	SMPL
J4583-06 - 1	TOD1-EXC-2-4	1200	0.2588	-	7471B	PPB	22 Aug 2018 17:16:12	U	SMPL
J4585-01 - 1	WS-GRAVEL-STAIN	38895	8.6673	-	7471B	PPB	22 Aug 2018 17:18:30	U	SMPL
CCV64 - 1	CCV64	22349	4.9765	-	7471B	PPB	22 Aug 2018 17:20:46	U	SMPL
CCB64 - 1	CCB64	28	-0.0026	-	7471B	PPB	22 Aug 2018 17:23:05	U	SMPL
J4583-02LX5 - 1		152	0.0250	-	7471B	PPB	22 Aug 2018 17:25:21	U	SMPL
J4583-02A - 1		17821	3.9664	-	7471B	PPB	22 Aug 2018 17:27:37	U	SMPL
J4580-01DLX5 - 1		26747	5.9575	-	7471B	PPB	22 Aug 2018 17:29:54	U	SMPL
CCV65 - 1	CCV65	22391	4.9858	-	7471B	PPB	22 Aug 2018 17:32:10	U	SMPL
CCB65 - 1	CCB65	38	-0.0004	-	7471B	PPB	22 Aug 2018 17:34:26	U	SMPL