



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB97485

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	08/22/2018	17:10
2	CAL2	1	Water			21.9	7.01	08/22/2018	17:12
3	CAL3	1	Water			22.2	10.04	08/22/2018	17:14
4	ICV	1	Water			21.8	7.00	08/22/2018	17:16
5	CCV1	1	Water			21.9	2.00	08/22/2018	17:18
6	J4295-01	1	Water			18.2	7.24	08/22/2018	17:20
7	J4295-01DUP	1	Water			18.3	7.25	08/22/2018	17:22
8	J4609-01	1	Water			24.7	8.51	08/22/2018	17:24
9	J4612-07	1	Water			28.1	7.06	08/22/2018	17:28
10	CCV2	1	Water			21.7	12.00	08/22/2018	17:30

WORKLIST(Hardcopy Internal Chain)

LB 97485

WorkList Name : PH082218

WorkList ID : 115939

Date : 8/22/2018 2:48:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2018	Water	J4295-01	pH	Cool 4 deg C	MINO01	E52	OUTFALL	08/21/2018	9040C
08/22/2018	Water	J4609-01	pH	Cool 4 deg C	PSEG01	E11	2749	08/22/2018	9040C
08/22/2018	Water	J4612-07	pH	Cool 4 deg C	PSEG01	E11	TOTE	08/22/2018	9040C

Date/Time 08/22/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 08/22/18 6:00
 Received by: CP
 Relinquished by: SP