



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : AMANDEEP

Run Number: LB97487

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	08/22/2018	17:10
2	CAL2	1	Water			21.9	7.01	08/22/2018	17:12
3	CAL3	1	Water			22.2	10.04	08/22/2018	17:14
4	ICV	1	Water			21.8	7.00	08/22/2018	17:16
5	CCV1	1	Water			22.0	2.00	08/22/2018	17:40
6	J4584-01	1	Solid	20.02	20	22.4	7.31	08/22/2018	17:42
7	J4584-01DUP	1	Solid	20.03	20	22.5	7.32	08/22/2018	17:44
8	J4594-04	1	Solid	20.02	20	24.0	9.26	08/22/2018	17:46
9	CCV2	1	Water			21.9	12.00	08/22/2018	17:54

WORKLIST(Hardcopy Internal Chain)

LB 97487

WorkList Name : corrosivity082218

WorkList ID : 115944

Date : 8/22/2018 3:35:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/21/2018	Solid	J4584-01	Corrosivity	Cool 4 deg C	PARS15	E42	TOD1-WC-S-1-0-52	08/21/2018	9040C
08/22/2018	Solid	J4594-04	Corrosivity	Cool 4 deg C	PSEG01	E11	1-3-RO	08/22/2018	9040C

Date/Time 08/22/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 08/22/18 6:00
 Received by: CP
 Relinquished by: SP