



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97488

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	08/22/2018	17:10
2	CAL2	1	Water			21.9	7.01	08/22/2018	17:12
3	CAL3	1	Water			22.2	10.04	08/22/2018	17:14
4	ICV	1	Water			21.8	7.00	08/22/2018	17:16
5	CCV1	1	Water			22.0	2.00	08/22/2018	17:40
6	J4612-03	1	Solid	20.03	20	24.7	7.41	08/22/2018	17:48
7	J4612-03DUP	1	Solid	20.04	20	24.8	7.42	08/22/2018	17:50
8	J4612-06	1	Solid	20.04	20	24.4	7.25	08/22/2018	17:52
9	CCV2	1	Water			21.9	12.00	08/22/2018	17:54

WORKLIST(Hardcopy Internal Chain)

LB 97488

WorkList Name : corrosion082218-2

WorkList ID : 115970

Date : 8/23/2018 3:35:50 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/22/2018	Solid	J4612-03	Corrosivity	Cool 4 deg C	PSEG01	E11	RBR200028	08/22/2018	9045D
08/22/2018	Solid	J4612-06	Corrosivity	Cool 4 deg C	PSEG01	E11	RBR200027	08/22/2018	9045D

Date/Time 08/22/18 5:00
 Received by: SP
 Relinquished by: CP

Date/Time 08/22/18 6:00
 Received by: CP
 Relinquished by: SP