


7470A

LB97489
INSTRUMENT ID: CVJ

Linear

 μ Abs.:

36081



A= 0.0000e+000

B= 2.7337e-004

C= -3.6800e-003

Rho= 0.9995221

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.003	-0.003	4	0.000	4				
0.2	0.200	0.232	0.032	861	0.0 %	861				
2.5	2.500	2.351	-0.149	8612	0.0 %	8612				
5.0	5.000	5.105	0.105	18688	0.0 %	18688				
7.5	7.500	7.655	0.155	28016	0.0 %	28016				
10.0	10.000	9.860	-0.140	36081	0.0 %	36081				

LB97489 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 23 Aug 2018 10:15:13

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	4	-	0.0000	7470A	PPB	23 Aug 2018 10:18:36	S	Std
0.2 - 1	50.2	861	-	0.2000	7470A	PPB	23 Aug 2018 10:20:44	S	Std
2.5 - 1	52.5	8612	-	2.5000	7470A	PPB	23 Aug 2018 10:22:53	S	Std
5.0 - 1	55	18688	-	5.0000	7470A	PPB	23 Aug 2018 10:25:01	S	Std
7.5 - 1	57.5	28016	-	7.5000	7470A	PPB	23 Aug 2018 10:27:08	S	Std
10.0 - 1	59	36081	-	10.0000	7470A	PPB	23 Aug 2018 10:31:45	S	Std
ICV25 - 1	ICV25	14470	3.9520	-7470A	PPB	23 Aug 2018 10:34:20	U	SMPL	
ICB25 - 1	ICB25	-209	-0.0608	-7470A	PPB	23 Aug 2018 10:36:26	U	SMPL	
CCV66 - 1	CCV66	19026	5.1975	-7470A	PPB	23 Aug 2018 10:38:34	U	SMPL	
CCB66 - 1	CCB66	-195	-0.0570	-7470A	PPB	23 Aug 2018 10:40:41	U	SMPL	
CRA - 1	CRA	838	0.2254	-7470A	PPB	23 Aug 2018 10:42:47	U	SMPL	
HighStd - 1	HighStd	36103	9.8660	-7470A	PPB	23 Aug 2018 10:44:54	U	SMPL	
ChkStd - 1	ChkStd	26354	7.2008	-7470A	PPB	23 Aug 2018 10:47:01	U	SMPL	
PB112298BL - 1	PBW	-16	-0.0081	-7470A	PPB	23 Aug 2018 10:52:19	U	SMPL	
PB112298BS - 1	LCSW	13884	3.7919	-7470A	PPB	23 Aug 2018 10:54:26	U	SMPL	
PB112298TB - 1	PB112268TB	-127	-0.0384	-7470A	PPB	23 Aug 2018 10:56:34	U	SMPL	
J4276-04 - 1	OK-01-082118	-20	-0.0091	-7470A	PPB	23 Aug 2018 10:58:43	U	SMPL	
J4278-08 - 1	EO-01-082118-A	3	-0.0029	-7470A	PPB	23 Aug 2018 11:00:51	U	SMPL	
J4278-10 - 1	EO-01-082118-B	4	-0.0026	-7470A	PPB	23 Aug 2018 11:03:00	U	SMPL	
J4278-12 - 1	EO-01-082118-C	2	-0.0031	-7470A	PPB	23 Aug 2018 11:05:06	U	SMPL	
CCV67 - 1	CCV67	18819	5.1410	-7470A	PPB	23 Aug 2018 11:07:13	U	SMPL	
CCB67 - 1	CCB67	-191	-0.0559	-7470A	PPB	23 Aug 2018 11:09:20	U	SMPL	
J4283-04 - 1	TR-05-082118	-22	-0.0097	-7470A	PPB	23 Aug 2018 11:11:28	U	SMPL	
J4283-06 - 1	TR-04-082118	-3	-0.0045	-7470A	PPB	23 Aug 2018 11:13:35	U	SMPL	
J4580-04 - 1	RN-SOIL	39	0.0070	-7470A	PPB	23 Aug 2018 11:15:42	U	SMPL	
J4585-02 - 1	WS-GRAVEL-STAIN	27	0.0037	-7470A	PPB	23 Aug 2018 11:17:49	U	SMPL	
J4585-02DUP - 1	WS-GRAVEL-STAIN DUP	6	-0.0020	-7470A	PPB	23 Aug 2018 11:19:56	U	SMPL	
J4585-02MS - 1	WS-GRAVEL-STAINMS	13798	3.7683	-7470A	PPB	23 Aug 2018 11:22:03	U	SMPL	
J4585-02MSD - 1	WS-GRAVEL-STAINMSD	14026	3.8307	-7470A	PPB	23 Aug 2018 11:24:11	U	SMPL	
PB112297BL - 1	PBW	-162	-0.0480	-7470A	PPB	23 Aug 2018 11:26:20	U	SMPL	
PB112297BS - 1	LCSW	13476	3.6803	-7470A	PPB	23 Aug 2018 11:28:29	U	SMPL	
PB112297TB - 1	PB112265TB	-134	-0.0403	-7470A	PPB	23 Aug 2018 11:30:36	U	SMPL	
CCV68 - 1	CCV68	19211	5.2481	-7470A	PPB	23 Aug 2018 11:32:42	U	SMPL	
CCB68 - 1	CCB68	-187	-0.0548	-7470A	PPB	23 Aug 2018 11:34:49	U	SMPL	
J4556-05 - 1	20180533-AMENDED-COMPOSITE-(ABC)	739	0.1983	-7470A	PPB	23 Aug 2018 11:36:58	U	SMPL	
J4556-05DUP - 1	20180533-AMENDED-COMPOSITE-(ABC)DUP	746	0.2003	-7470A	PPB	23 Aug 2018 11:39:05	U	SMPL	
J4556-05MS - 1	20180533-AMENDED-COMPOSITE-(ABC)MS	14829	4.0502	-7470A	PPB	23 Aug 2018 11:41:12	U	SMPL	
J4556-05MSD - 1	20180533-AMENDED-COMPOSITE-(ABC)MSD	14836	4.0521	-7470A	PPB	23 Aug 2018 11:43:19	U	SMPL	
J4556-06 - 1	20180534-AMENDED-COMPOSITE-(DEF)	269	0.0699	-7470A	PPB	23 Aug 2018 11:45:27	U	SMPL	
J4556-07 - 1	20180535-AMENDED-COMPOSITE-(GHI)	519	0.1382	-7470A	PPB	23 Aug 2018 11:47:35	U	SMPL	
J4585-02LX5 - 1		-3	-0.0045	-7470A	PPB	23 Aug 2018 11:49:44	U	SMPL	
J4585-02A - 1		13705	3.7429	-7470A	PPB	23 Aug 2018 11:51:52	U	SMPL	
J4556-05LX5 - 1		14	0.0001	-7470A	PPB	23 Aug 2018 11:54:00	U	SMPL	
J4556-05A - 1		15228	4.1593	-7470A	PPB	23 Aug 2018 11:56:07	U	SMPL	
CCV69 - 1	CCV69	19288	5.2692	-7470A	PPB	23 Aug 2018 11:58:15	U	SMPL	
CCB69 - 1	CCB69	-197	-0.0575	-7470A	PPB	23 Aug 2018 12:00:22	U	SMPL	