



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB97495

Ambient Barometric Pressure (mmHg): 755

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.9	08/23/2018	09:40
2	J4609-02		1	8	>100.0	>212.0	>212.0	08/23/2018	10:05
3	J4609-02DUP		1	8	>100.0	>212.0	>212.0	08/23/2018	10:40
4	J4612-08		1	6	>100.0	>212.0	>212.0	08/23/2018	11:10

$$\text{Result} = (\text{Celsius} * 1.8) + 32$$

$$\text{Final Result} = \text{Result} + (760 - \text{Ambient Barometric Pressure}) * 0.06$$