



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB97500

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP67267
Turbidity std, 40NTU	WP67262
turbidity std 5 NTU	WP67265
turbidity std 20 NTU	WP67263
Turbidity std 1 NTU	WP67266
turbidity std 0 NTU	WP67268
turbidity std 10 NTU	WP67264
10 NTU Standard 500 ml	W2297

Intercept: 0.2076

Slope: 0.9216

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.235	08/23/2018	12:57
2	CAL2	1	1	0.992	08/23/2018	12:59
3	CAL3	5	1	4.900	08/23/2018	13:02
4	CAL4	10	1	9.760	08/23/2018	13:07
5	CAL5	20	1	18.200	08/23/2018	13:10
6	CAL6	40	1	37.200	08/23/2018	13:14



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB97500

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.400	9.974	08/23/2018	13:17
2	ICB		1	0.230	0.024	08/23/2018	13:20
3	CCV1	10	1	9.560	10.148	08/23/2018	13:23
4	CCB1		1	0.236	0.031	08/23/2018	13:26
5	LB97500BL		1	0.231	0.025	08/23/2018	13:29
6	J4619-01		1	26.300	28.312	08/23/2018	13:33
7	J4619-01DUP		1	25.200	27.118	08/23/2018	13:36
8	CCV2	10	1	9.650	10.246	08/23/2018	13:40
9	CCB2		1	0.231	0.025	08/23/2018	13:43

UB97500

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TURBIDITY082318

WorkList ID : 115979

Date : 8/23/2018 11:13:10 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/24/2018	Water	J4619-01	Turbidity	Cool 4 deg C	LIRO01	E31	UDS-180822-A	08/22/2018	SM2130 B

Date/Time 8/23/18 12:30

Received by: AS

Relinquished by: AS

Date/Time 8/23/18 13:50

Received by: CP

Relinquished by: AR