

LB 97501


245.1

INSTRUMENT ID: CV2

Linear

μ Abs.:

36378



A= 0.0000e+000

B= 2.7666e-004

C= 4.1558e-003

Rho= 0.9997624

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	38	0.000	38				
0.05	0.050	0.047	-0.003	154	0.0 %	154				
0.2	0.200	0.184	-0.016	649	0.0 %	649				
2.5	2.500	2.495	-0.005	9003	0.0 %	9003				
5.0	5.000	5.110	0.110	18456	0.0 %	18456				
7.5	7.500	7.331	-0.169	26484	0.0 %	26484				
10.0	10.000	10.069	0.069	36378	0.0 %	36378				

LB97501 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 23 Aug 2018 13:53:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	51	38	-	0.0000	245.1	PPB	23 Aug 2018 13:57:40	S	Std
0.05 - 1	50.05	154	-	0.0500	245.1	PPB	23 Aug 2018 13:59:58	S	Std
0.2 - 1	50.2	649	-	0.2000	245.1	PPB	23 Aug 2018 14:02:15	S	Std
2.5 - 1	52.5	9003	-	2.5000	245.1	PPB	23 Aug 2018 14:04:33	S	Std
5.0 - 1	55	18456	-	5.0000	245.1	PPB	23 Aug 2018 14:06:50	S	Std
7.5 - 1	57.5	26484	-	7.5000	245.1	PPB	23 Aug 2018 14:09:06	S	Std
10.0 - 1	510	36378	-	10.0000	245.1	PPB	23 Aug 2018 14:13:59	S	Std
ICV26 - 1	ICV26	14417	3.9928	-245.1		PPB	23 Aug 2018 14:16:54	U	SMPL
ICB26 - 1	ICB26	-230	-0.0595	-245.1		PPB	23 Aug 2018 14:19:09	U	SMPL
CCV70 - 1	CCV70	18381	5.0894	-245.1		PPB	23 Aug 2018 14:21:26	U	SMPL
CCB70 - 1	CCB70	-206	-0.0528	-245.1		PPB	23 Aug 2018 14:23:42	U	SMPL
CRA - 1	CRA	128	0.0396	-245.1		PPB	23 Aug 2018 14:25:59	U	SMPL
HighStd - 1	HighStd	36220	10.0248	-245.1		PPB	23 Aug 2018 14:28:15	U	SMPL
ChkStd - 1	ChkStd	24017	6.6487	-245.1		PPB	23 Aug 2018 14:30:31	U	SMPL
PB112304BL - 1	PBW	-15	0.0000	-245.1		PPB	23 Aug 2018 14:35:32	U	SMPL
PB112304BS - 1	LCSW	13927	3.8572	-245.1		PPB	23 Aug 2018 14:37:49	U	SMPL
J4568-01 - 1	081718-SIDI-E-D-B	-117	-0.0282	-245.1		PPB	23 Aug 2018 14:40:05	U	SMPL
J4568-03 - 1	081718-SIDI-E-U-B	-23	-0.0022	-245.1		PPB	23 Aug 2018 14:42:24	U	SMPL
J4568-04 - 1	081718-SIDI-E-U-S	-30	-0.0041	-245.1		PPB	23 Aug 2018 14:44:42	U	SMPL
J4568-06 - 1	081718-DBRI-E-D-S	-27	-0.0033	-245.1		PPB	23 Aug 2018 14:47:00	U	SMPL
J4568-07 - 1	081718-DBRI-E-U-B	-5	0.0028	-245.1		PPB	23 Aug 2018 14:49:15	U	SMPL
CCV71 - 1	CCV71	18499	5.1221	-245.1		PPB	23 Aug 2018 14:51:31	U	SMPL
CCB71 - 1	CCB71	-209	-0.0537	-245.1		PPB	23 Aug 2018 14:53:47	U	SMPL
J4582-01 - 1	082018-DBRI-E-D-B	-34	-0.0053	-245.1		PPB	23 Aug 2018 14:56:04	U	SMPL
J4582-02 - 1	082018-DBRI-E-D-M	-34	-0.0053	-245.1		PPB	23 Aug 2018 14:58:20	U	SMPL
J4582-03 - 1	082018-DBRI-E-D-S	-18	-0.0008	-245.1		PPB	23 Aug 2018 15:00:37	U	SMPL
J4582-04 - 1	082018-DBRI-E-U-B	-3	0.0033	-245.1		PPB	23 Aug 2018 15:02:53	U	SMPL
J4582-05 - 1	082018-DBRI-E-U-M	-19	-0.0011	-245.1		PPB	23 Aug 2018 15:05:09	U	SMPL
J4582-06 - 1	082018-DBRI-E-U-S	-17	-0.0005	-245.1		PPB	23 Aug 2018 15:07:26	U	SMPL
J4582-07 - 1	082018-TIPC-E-D-M	-5	0.0028	-245.1		PPB	23 Aug 2018 15:09:43	U	SMPL
J4582-08 - 1	082018-TIPC-E-U-M	-11	0.0011	-245.1		PPB	23 Aug 2018 15:12:01	U	SMPL
J4582-09 - 1	082018-SIPI-E-D-B	-23	-0.0022	-245.1		PPB	23 Aug 2018 15:14:18	U	SMPL
J4582-10 - 1	082018-SIPI-E-D-S	-13	0.0006	-245.1		PPB	23 Aug 2018 15:16:34	U	SMPL
CCV72 - 1	CCV72	18409	5.0972	-245.1		PPB	23 Aug 2018 15:18:51	U	SMPL
CCB72 - 1	CCB72	-190	-0.0484	-245.1		PPB	23 Aug 2018 15:21:07	U	SMPL
J4582-11 - 1	082018-SIPI-E-U-B	-62	-0.0130	-245.1		PPB	23 Aug 2018 15:23:24	U	SMPL
J4582-12 - 1	082018-SIPI-E-U-S	-22	-0.0019	-245.1		PPB	23 Aug 2018 15:25:40	U	SMPL
J4582-12DUP - 1	082018-SIPI-E-U-SDUP	-27	-0.0033	-245.1		PPB	23 Aug 2018 15:27:56	U	SMPL
J4582-12MS - 1	082018-SIPI-E-U-SMS	15000	4.1541	-245.1		PPB	23 Aug 2018 15:30:13	U	SMPL
J4582-12MSD - 1	082018-SIPI-E-U-SMSD	14981	4.1488	-245.1		PPB	23 Aug 2018 15:32:30	U	SMPL
J4582-12LX5 - 1		-141	-0.0349	-245.1		PPB	23 Aug 2018 15:34:47	U	SMPL
J4582-12A - 1		15709	4.3502	-245.1		PPB	23 Aug 2018 15:37:05	U	SMPL
CCV73 - 1	CCV73	18370	5.0864	-245.1		PPB	23 Aug 2018 15:39:22	U	SMPL
CCB73 - 1	CCB73	-228	-0.0589	-245.1		PPB	23 Aug 2018 15:41:41	U	SMPL