



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By :

Run Number: LB97520

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/23/2018	17:00
2	CAL2	1	Water			22.2	7.01	08/23/2018	17:02
3	CAL3	1	Water			22.1	10.04	08/23/2018	17:04
4	ICV	1	Water			22.0	7.00	08/23/2018	17:06
5	CCV1	1	Water			21.8	2.00	08/23/2018	17:16
6	J4273-02	1	Solid	20.02	20	23.9	8.23	08/23/2018	17:18
7	J4273-02DUP	1	Solid	20.03	20	24.0	8.24	08/23/2018	17:20
8	J4620-01	1	Solid	20.02	20	22.9	5.49	08/23/2018	17:22
9	J4620-02	1	Solid	20.02	20	23.0	5.31	08/23/2018	17:24
10	J4620-03	1	Solid	20.02	20	22.8	5.58	08/23/2018	17:26
11	J4621-05	1	Solid	20.04	20	21.6	8.28	08/23/2018	17:28
12	J4621-09	1	Solid	20.02	20	21.2	8.32	08/23/2018	17:30
13	CCV2	1	Water			22.0	12.00	08/23/2018	17:40

WORKLIST(Hardcopy Internal Chain)

LB97520

WorkList Name : corrosivity082318

WorkList ID : 115993

Date : 8/23/2018 4:36:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/23/2018	Solid	J4273-02	Corrosivity	Cool 4 deg C	PSEG05	E22	NB-07-082318	08/23/2018	9045D
08/23/2018	Solid	J4620-01	Corrosivity	Cool 4 deg C	PSEG01	E21	DRUM-1	08/23/2018	9045D
08/23/2018	Solid	J4620-02	Corrosivity	Cool 4 deg C	PSEG01	E21	DRUM-2	08/23/2018	9045D
08/23/2018	Solid	J4620-03	Corrosivity	Cool 4 deg C	PSEG01	E21	DRUM-3	08/23/2018	9045D
08/23/2018	Solid	J4621-05	Corrosivity	Cool 4 deg C	PSEG01	E22	SP-1	08/23/2018	9045D
08/23/2018	Solid	J4621-09	Corrosivity	Cool 4 deg C	PSEG01	E22	SP-2	08/23/2018	9045D

Date/Time 08/23/18 4:40

Received by: SP

Relinquished by: CP

Date/Time 08/23/18 6:15

Received by: CP

Relinquished by: SP