



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : JIGNESH

Run Number: LB97521

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/23/2018	17:00
2	CAL2	1	Water			22.2	7.01	08/23/2018	17:02
3	CAL3	1	Water			22.1	10.04	08/23/2018	17:04
4	ICV	1	Water			22.0	7.00	08/23/2018	17:06
5	CCV1	1	Water			21.8	2.00	08/23/2018	17:16
6	J4270-02	1	Solid	20.02	20	24.4	5.73	08/23/2018	17:32
7	J4270-02DUP	1	Solid	20.03	20	24.5	5.74	08/23/2018	17:34
8	J4270-04	1	Solid	20.02	20	24.1	5.65	08/23/2018	17:36
9	J4282-12	1	Solid	20.04	20	23.8	8.42	08/23/2018	17:38
10	J4282-14	1	Solid	20.02	20	22.7	9.31	08/23/2018	17:40
11	CCV2	1	Water			22.0	12.00	08/23/2018	17:42

WORKLIST(Hardcopy Internal Chain)

LB 97521

WorkList Name : corrosivity082318-2

WorkList ID : 115997

Date : 8/23/2018 4:50:17 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/23/2018	Solid	J4270-02	Corrosivity	Cool 4 deg C	PSEG05	E31	BU-03-082318	08/23/2018	9040C
08/23/2018	Solid	J4270-04	Corrosivity	Cool 4 deg C	PSEG05	E31	BU-01-082318	08/23/2018	9040C
08/23/2018	Solid	J4282-12	Corrosivity	Cool 4 deg C	PSEG05		SU-02-082318-A	08/23/2018	9040C
08/23/2018	Solid	J4282-14	Corrosivity	Cool 4 deg C	PSEG05		SU-02-082318-B	08/23/2018	9040C

Date/Time 08/23/18 4:40

Received by: SP

Relinquished by: CP

Date/Time 08/23/18 6:15

Received by: CP

Relinquished by: SP