



Analytical Summary Report

Analysis Method: 9040C

Analyst By : sweta

Parameter: pH

Supervisor Review By : JIGNESH

Run Number: LB97524

Slope : 99.0

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/23/2018	17:00
2	CAL2	1	Water			22.2	7.01	08/23/2018	17:02
3	CAL3	1	Water			22.1	10.04	08/23/2018	17:04
4	ICV	1	Water			22.0	7.00	08/23/2018	17:06
5	CCV1	1	Water			21.9	2.00	08/23/2018	17:50
6	J4616-01	1	Water			23.2	7.46	08/23/2018	17:52
7	J4616-01DUP	1	Water			23.3	7.47	08/23/2018	17:54
8	J4617-01	1	Water			23.4	7.68	08/23/2018	17:56
9	J4617-04	1	Water			22.9	7.56	08/23/2018	17:58
10	J4621-10	1	Water			23.6	9.96	08/23/2018	18:00
11	J4627-01	1	Water			23.4	7.82	08/23/2018	18:02
12	J4638-01	1	Water			22.0	7.65	08/23/2018	18:04
13	J4642-01	1	Water			25.7	7.29	08/23/2018	18:06
14	CCV2	1	Water			21.9	12.00	08/23/2018	18:08

WORKLIST(Hardcopy Internal Chain)

LB 97524

WorkList Name : ph082318-1

WorkList ID : 115996

Date : 8/23/2018 4:37:21 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/23/2018	Water	J4616-01	pH	Cool 4 deg C	PSEG01	E12	VB27184	08/23/2018	9040C
08/23/2018	Water	J4617-01	pH	Cool 4 deg C	PSEG01	E12	N48969	08/23/2018	9040C
08/23/2018	Water	J4617-04	pH	Cool 4 deg C	PSEG01	E12	N1332	08/23/2018	9040C
08/23/2018	Water	J4621-10	pH	Cool 4 deg C	PSEG01	E22	OILY-WATER	08/23/2018	9040C
08/23/2018	Water	J4627-01	pH	Cool 4 deg C	PSEG01	E22	VB-N29937	08/23/2018	9040C
08/23/2018	Water	J4638-01	pH	Cool 4 deg C	PSEG01	E31	PS-N48964	08/23/2018	9040C
08/23/2018	Water	J4642-01	pH	Cool 4 deg C	PSEG01	E22	2727	08/23/2018	9040C

Date/Time 08/23/18 4:40

Received by: SP

Relinquished by: CP

Date/Time 08/23/18 6:15

Received by: CP

Relinquished by: SP