

7470A


LB 97526

INSTRUMENT ID: CV2

Linear

 μ Abs.:

36048



A= 0.0000e+000

B= 2.7506e-004

C= 1.2646e-002

Rho= 0.9998659

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	17	0.000	17				
0.2	0.200	0.202	0.002	689	0.0 %	689				
2.5	2.500	2.422	-0.078	8759	0.0 %	8759				
5.0	5.000	5.035	0.035	18258	0.0 %	18258				
7.5	7.500	7.596	0.096	27568	0.0 %	27568				
10.0	10.000	9.928	-0.072	36048	0.0 %	36048				

LB97526 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Aug 2018 10:14:22

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	17	-	0.00007470A	PPB	24 Aug 2018 10:17:22	S	Std
0.2 - 1	50.2	689	-	0.20007470A	PPB	24 Aug 2018 10:19:30	S	Std
2.5 - 1	52.5	8759	-	2.50007470A	PPB	24 Aug 2018 10:21:38	S	Std
5.0 - 1	55	18258	-	5.00007470A	PPB	24 Aug 2018 10:23:46	S	Std
7.5 - 1	57.5	27568	-	7.50007470A	PPB	24 Aug 2018 10:25:54	S	Std
10.0 - 1	510	36048	-	10.00007470A	PPB	24 Aug 2018 10:30:37	S	Std
ICV28 - 1	ICV28	14202	3.9191	-7470A	PPB	24 Aug 2018 10:33:09	U	SMPL
ICB28 - 1	ICB28	-197	-0.0415	-7470A	PPB	24 Aug 2018 10:35:16	U	SMPL
CCV77 - 1	CCV77	17629	4.8618	-7470A	PPB	24 Aug 2018 10:37:24	U	SMPL
CCB77 - 1	CCB77	-149	-0.0283	-7470A	PPB	24 Aug 2018 10:39:30	U	SMPL
CRA - 1	CRA	656	0.1931	-7470A	PPB	24 Aug 2018 10:41:38	U	SMPL
HighStd - 1	HighStd	36231	9.9785	-7470A	PPB	24 Aug 2018 10:43:44	U	SMPL
ChkStd - 1	ChkStd	25449	7.0128	-7470A	PPB	24 Aug 2018 10:45:51	U	SMPL
PB112330BL - 1	PBW	-43	0.0008	-7470A	PPB	24 Aug 2018 10:50:28	U	SMPL
PB112330BS - 1	LCSW	13386	3.6947	-7470A	PPB	24 Aug 2018 10:52:36	U	SMPL
PB112330TB - 1	PB112300TB	-151	-0.0289	-7470A	PPB	24 Aug 2018 10:54:44	U	SMPL
J4584-01 - 1	TOD1-WC-S-1-0-52	-23	0.0063	-7470A	PPB	24 Aug 2018 10:56:52	U	SMPL
J4584-01DUP - 1	TOD1-WC-S-1-0-52DUP	-31	0.0041	-7470A	PPB	24 Aug 2018 10:59:00	U	SMPL
J4584-01MS - 1	TOD1-WC-S-1-0-52MS	12938	3.5714	-7470A	PPB	24 Aug 2018 11:01:08	U	SMPL
J4584-01MSD - 1	TOD1-WC-S-1-0-52MSD	14062	3.8806	-7470A	PPB	24 Aug 2018 11:03:15	U	SMPL
CCV78 - 1	CCV78	17735	4.8909	-7470A	PPB	24 Aug 2018 11:05:22	U	SMPL
CCB78 - 1	CCB78	-172	-0.0347	-7470A	PPB	24 Aug 2018 11:07:29	U	SMPL
J4594-04 - 1	1-3-RO	132	0.0490	-7470A	PPB	24 Aug 2018 11:09:37	U	SMPL
J4612-03 - 1	RBR200028	-53	-0.0019	-7470A	PPB	24 Aug 2018 11:11:44	U	SMPL
J4612-06 - 1	RBR200027	-28	0.0049	-7470A	PPB	24 Aug 2018 11:13:51	U	SMPL
PB112331BL - 1	PBW	-37	0.0025	-7470A	PPB	24 Aug 2018 11:15:58	U	SMPL
PB112331BS - 1	LCSW	14301	3.9463	-7470A	PPB	24 Aug 2018 11:18:06	U	SMPL
PB112331TB - 1	PB112303TB	-127	-0.0223	-7470A	PPB	24 Aug 2018 11:20:13	U	SMPL
J4609-02 - 1	2749	18	0.0176	-7470A	PPB	24 Aug 2018 11:22:21	U	SMPL
J4609-02DUP - 1	2749DUP	31	0.0212	-7470A	PPB	24 Aug 2018 11:24:29	U	SMPL
J4609-02MS - 1	2749MS	14680	4.0506	-7470A	PPB	24 Aug 2018 11:26:38	U	SMPL
J4609-02MSD - 1	2749MSD	14305	3.9474	-7470A	PPB	24 Aug 2018 11:28:45	U	SMPL
CCV79 - 1	CCV79	17400	4.7988	-7470A	PPB	24 Aug 2018 11:30:52	U	SMPL
CCB79 - 1	CCB79	-186	-0.0385	-7470A	PPB	24 Aug 2018 11:33:00	U	SMPL
J4612-08 - 1	TOTE	136	0.0501	-7470A	PPB	24 Aug 2018 11:35:07	U	SMPL
PB112332BL - 1	PBW	5	0.0140	-7470A	PPB	24 Aug 2018 11:37:14	U	SMPL
PB112332BS - 1	LCSW	12916	3.5654	-7470A	PPB	24 Aug 2018 11:39:21	U	SMPL
J4609-01 - 1	2749	157	0.0558	-7470A	PPB	24 Aug 2018 11:41:28	U	SMPL
J4609-01DUP - 1	2749DUP	125	0.0470	-7470A	PPB	24 Aug 2018 11:43:36	U	SMPL
J4609-01MS - 1	2749MS	15203	4.1945	-7470A	PPB	24 Aug 2018 11:45:43	U	SMPL
J4609-01MSD - 1	2749MSD	15197	4.1928	-7470A	PPB	24 Aug 2018 11:47:51	U	SMPL
J4612-07 - 1	TOTE	27473	7.5695	-7470A	PPB	24 Aug 2018 11:50:00	U	SMPL
CCV80 - 1	CCV80	16961	4.6780	-7470A	PPB	24 Aug 2018 11:52:09	U	SMPL
CCB80 - 1	CCB80	-104	-0.0160	-7470A	PPB	24 Aug 2018 11:54:19	U	SMPL
J4584-01LX5 - 1		30	0.0209	-7470A	PPB	24 Aug 2018 11:56:27	U	SMPL
J4584-01A - 1		14191	3.9161	-7470A	PPB	24 Aug 2018 11:58:35	U	SMPL
J4609-02LX5 - 1		-69	-0.0063	-7470A	PPB	24 Aug 2018 12:00:42	U	SMPL
J4609-02A - 1		14499	4.0008	-7470A	PPB	24 Aug 2018 12:02:49	U	SMPL
J4609-01LX5 - 1		-91	-0.0124	-7470A	PPB	24 Aug 2018 12:04:56	U	SMPL
J4609-01A - 1		16970	4.6805	-7470A	PPB	24 Aug 2018 12:07:04	U	SMPL
CCV81 - 1	CCV81	18058	4.9798	-7470A	PPB	24 Aug 2018 12:09:11	U	SMPL
CCB81 - 1	CCB81	-181	-0.0371	-7470A	PPB	24 Aug 2018 12:11:19	U	SMPL