



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: AMANDEEP

ANALYST: Aparna

Date: 08/23/2018

TEMP1 IN: 103 08/23/2018 14:10 TEMP OUT: 104 08/23/2018 15:20

TEMP2 IN: 105 08/23/2018 16:00 TEMP OUT: 104 08/24/2018 07:00

TEMP3 IN: 105 08/24/2018 07:45 TEMP OUT: 105 08/24/2018 09:20

TEMP4 IN: 105 08/24/2018 10:00 TEMP OUT: 105 08/24/2018 11:35

Run Number: LB97530

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97530BL	LB97530BL	1.4025	1.4025	100	1.4025	1.4025	1.4025	0.0000	0
2	LB97530BS	LB97530BS	1.3926	1.3926	100	1.4430	1.4430	1.4430	0.0504	504
3	J4618-01	FM1026	1.3939	1.3939	900	1.3972	1.3972	1.3972	0.0033	3.7

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS082418-2

WorkList ID : 116037

Date : 8/24/2018 07:00:54 AM

2197530

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/28/2018	Water	J4618-01	TSS	Cool 4 deg C	PSEG01	K63	FM1026	08/23/2018	SM2540 D

Date/Time

Received by:

Relinquished by:

08/24/18, 07:15
AS
CP

Date/Time

Received by:

Relinquished by:

08/24/18, 09:30
AS
AS