


L397533

245.1

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 2.7780e-004

C= 1.2617e-002

Rho= 0.9999156

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.026	-0.026	-138	0.000	-138				
0.05	0.050	0.057	0.007	158	0.0 %	158				
0.2	0.200	0.248	0.048	848	0.0 %	848				
2.5	2.500	2.522	0.022	9032	0.0 %	9032				
5.0	5.000	4.951	-0.049	17778	0.0 %	17778				
7.5	7.500	7.427	-0.073	26689	0.0 %	26689				
10.0	10.000	10.071	0.071	36208	0.0 %	36208				

LB97533 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 24 Aug 2018 13:48:25

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-138	-	0.0000	245.1	PPB	24 Aug 2018 14:12:39	S	Std
0.05 - 1	50.05	158	-	0.0500	245.1	PPB	24 Aug 2018 14:14:56	S	Std
0.2 - 1	50.2	848	-	0.2000	245.1	PPB	24 Aug 2018 14:17:14	S	Std
2.5 - 1	52.5	9032	-	2.5000	245.1	PPB	24 Aug 2018 14:19:31	S	Std
5.0 - 1	55	17778	-	5.0000	245.1	PPB	24 Aug 2018 14:21:48	S	Std
7.5 - 1	57.5	26689	-	7.5000	245.1	PPB	24 Aug 2018 14:24:05	S	Std
10.0 - 1	510	36208	-	10.0000	245.1	PPB	24 Aug 2018 14:26:23	S	Std
ICV29 - 1	ICV29	14180	3.9518	-245.1		PPB	24 Aug 2018 14:29:21	U	SMPL
ICB29 - 1	ICB29	-50	-0.0013	-245.1		PPB	24 Aug 2018 14:31:36	U	SMPL
CCV82 - 1	CCV82	17961	5.0022	-245.1		PPB	24 Aug 2018 14:33:53	U	SMPL
CCB82 - 1	CCB82	-100	-0.0152	-245.1		PPB	24 Aug 2018 14:36:09	U	SMPL
CRA - 1	CRA	158	0.0565	-245.1		PPB	24 Aug 2018 14:38:26	U	SMPL
HighStd - 1	HighStd	36681	10.2026	-245.1		PPB	24 Aug 2018 14:40:42	U	SMPL
ChkStd - 1	ChkStd	25419	7.0740	-245.1		PPB	24 Aug 2018 14:42:58	U	SMPL
PB112333BL - 1	PBW	39	0.0235	-245.1		PPB	24 Aug 2018 14:47:58	U	SMPL
PB112333BS - 1	LCSW	13348	3.7207	-245.1		PPB	24 Aug 2018 14:50:14	U	SMPL
J4610-01 - 1	082118-DBRI-F-D-B	-149	-0.0288	-245.1		PPB	24 Aug 2018 14:52:31	U	SMPL
J4610-02 - 1	082118-DBRI-F-D-M	-49	-0.0010	-245.1		PPB	24 Aug 2018 14:54:50	U	SMPL
J4610-03 - 1	082118-DBRI-F-D-S	-25	0.0057	-245.1		PPB	24 Aug 2018 14:57:07	U	SMPL
J4610-04 - 1	082118-DBRI-F-U-B	8	0.0148	-245.1		PPB	24 Aug 2018 14:59:25	U	SMPL
J4610-05 - 1	082118-DBRI-F-U-M	-1	0.0123	-245.1		PPB	24 Aug 2018 15:01:41	U	SMPL
CCV83 - 1	CCV83	17576	4.8952	-245.1		PPB	24 Aug 2018 15:03:57	U	SMPL
CCB83 - 1	CCB83	-82	-0.0102	-245.1		PPB	24 Aug 2018 15:06:12	U	SMPL
J4610-06 - 1	082118-DBRI-F-U-S	-35	0.0029	-245.1		PPB	24 Aug 2018 15:08:30	U	SMPL
J4610-07 - 1	082118-SIDI-E-D-B	36	0.0226	-245.1		PPB	24 Aug 2018 15:10:45	U	SMPL
J4610-08 - 1	082118-SIDI-E-D-M	-12	0.0093	-245.1		PPB	24 Aug 2018 15:13:02	U	SMPL
J4610-09 - 1	082118-SIDI-E-D-S	-13	0.0090	-245.1		PPB	24 Aug 2018 15:15:19	U	SMPL
J4610-10 - 1	082118-SIDI-E-U-B	25	0.0196	-245.1		PPB	24 Aug 2018 15:17:35	U	SMPL
J4610-11 - 1	082118-SIDI-E-U-M	-4	0.0115	-245.1		PPB	24 Aug 2018 15:19:52	U	SMPL
J4610-12 - 1	082118-SIDI-E-U-S	-5	0.0112	-245.1		PPB	24 Aug 2018 15:22:09	U	SMPL
J4610-13 - 1	082118-TIPC-E-D-M	13	0.0162	-245.1		PPB	24 Aug 2018 15:24:26	U	SMPL
J4610-14 - 1	082118-TIPC-E-U-M	25	0.0196	-245.1		PPB	24 Aug 2018 15:26:45	U	SMPL
J4610-14DUP - 1	082118-TIPC-E-U-MDUP	2	0.0132	-245.1		PPB	24 Aug 2018 15:29:01	U	SMPL
CCV84 - 1	CCV84	17657	4.9177	-245.1		PPB	24 Aug 2018 15:31:17	U	SMPL
CCB84 - 1	CCB84	-89	-0.0121	-245.1		PPB	24 Aug 2018 15:33:34	U	SMPL
J4610-14MS - 1	082118-TIPC-E-U-MMS	14695	4.0949	-245.1		PPB	24 Aug 2018 15:35:51	U	SMPL
J4610-14MSD - 1	082118-TIPC-E-U-MMSD	14563	4.0582	-245.1		PPB	24 Aug 2018 15:38:08	U	SMPL
J4610-14LX5 - 1		-123	-0.0216	-245.1		PPB	24 Aug 2018 15:40:25	U	SMPL
J4610-14A - 1		16956	4.7230	-245.1		PPB	24 Aug 2018 15:42:41	U	SMPL
CCV85 - 1	CCV85	17561	4.8911	-245.1		PPB	24 Aug 2018 15:44:58	U	SMPL
CCB85 - 1	CCB85	-96	-0.0141	-245.1		PPB	24 Aug 2018 15:47:15	U	SMPL