


7471B

LB97542
INSTRUMENT ID: LY2

Linear

 μ Abs.:

44379



A= 0.0000e+000

B= 2.2464e-004

C= -2.7235e-002

Rho= 0.9999454

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.028	-0.028	-3	0.000	-3				
0.2	0.200	0.175	-0.025	902	0.0 %	902				
2.5	2.500	2.529	0.029	11381	0.0 %	11381				
5.0	5.000	5.042	0.042	22568	0.0 %	22568				
7.5	7.500	7.539	0.039	33681	0.0 %	33681				
10.0	10.000	9.942	-0.058	44379	0.0 %	44379				

LB97542 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Aug 2018 15:58:06

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-3	-	0.00007471B	PPB		24 Aug 2018 16:01:17	S	Std
0.2 - 1	50.2	902	-	0.20007471B	PPB		24 Aug 2018 16:03:35	S	Std
2.5 - 1	52.5	11381	-	2.50007471B	PPB		24 Aug 2018 16:05:52	S	Std
5.0 - 1	53	22568	-	5.00007471B	PPB		24 Aug 2018 16:08:09	S	Std
7.5 - 1	57.5	33681	-	7.50007471B	PPB		24 Aug 2018 16:10:27	S	Std
10.0 - 1	511	44379	-	10.00007471B	PPB		24 Aug 2018 16:18:58	S	Std
ICV30 - 1	ICV30	18411	4.1086	-7471B	PPB		24 Aug 2018 16:21:39	U	SMPL
ICB30 - 1	ICB30	-176	-0.0668	-7471B	PPB		24 Aug 2018 16:23:54	U	SMPL
CCV86 - 1	CCV86	22956	5.1296	-7471B	PPB		24 Aug 2018 16:26:10	U	SMPL
CCB86 - 1	CCB86	-162	-0.0636	-7471B	PPB		24 Aug 2018 16:28:26	U	SMPL
CRA - 1	CRA	868	0.1678	-7471B	PPB		24 Aug 2018 16:30:43	U	SMPL
HighStd - 1	HighStd	44896	10.0581	-7471B	PPB		24 Aug 2018 16:32:59	U	SMPL
ChkStd - 1	ChkStd	31525	7.0545	-7471B	PPB		24 Aug 2018 16:40:35	U	SMPL
PB112341BL - 1	PBS	12	-0.0245	-7471B	PPB		24 Aug 2018 16:45:25	U	SMPL
PB112341BS - 1	LCSS	15465	3.4468	-7471B	PPB		24 Aug 2018 16:48:54	U	SMPL
J4581-01 - 1	EP1-Q3-A1	3773	0.8203	-7471B	PPB		24 Aug 2018 16:51:11	U	SMPL
J4581-02 - 1	EP1-Q3-A2	1984	0.4184	-7471B	PPB		24 Aug 2018 16:53:28	U	SMPL
J4581-03 - 1	EP1-Q3-A3	939	0.1837	-7471B	PPB		24 Aug 2018 16:55:45	U	SMPL
J4581-04 - 1	EP1-Q3-A4	484	0.0815	-7471B	PPB		24 Aug 2018 16:58:03	U	SMPL
J4581-05 - 1	EP1-Q3-B1	1910	0.4018	-7471B	PPB		24 Aug 2018 17:00:18	U	SMPL
CCV87 - 1	CCV87	22624	5.0550	-7471B	PPB		24 Aug 2018 17:02:35	U	SMPL
CCB87 - 1	CCB87	-131	-0.0567	-7471B	PPB		24 Aug 2018 17:04:50	U	SMPL
J4581-06 - 1	EP1-Q3-B2	2941	0.6334	-7471B	PPB		24 Aug 2018 17:07:06	U	SMPL
J4581-07 - 1	EP1-Q3-B3	1637	0.3405	-7471B	PPB		24 Aug 2018 17:09:22	U	SMPL
J4581-08 - 1	EP1-Q3-B4	4228	0.9225	-7471B	PPB		24 Aug 2018 17:11:38	U	SMPL
J4581-09 - 1	EP1-Q3-C1	2478	0.5294	-7471B	PPB		24 Aug 2018 17:13:55	U	SMPL
J4581-10 - 1	EP1-Q3-C2	1587	0.3293	-7471B	PPB		24 Aug 2018 17:16:12	U	SMPL
J4581-10DUP - 1	EP1-Q3-C2DUP	1561	0.3234	-7471B	PPB		24 Aug 2018 17:18:28	U	SMPL
J4581-11 - 1	J4581-10MS	18813	4.1989	-7471B	PPB		24 Aug 2018 17:20:46	U	SMPL
J4581-12 - 1	J4581-10MSD	16557	3.6921	-7471B	PPB		24 Aug 2018 17:23:03	U	SMPL
J4581-13 - 1	EP1-Q3-C3	761	0.1437	-7471B	PPB		24 Aug 2018 17:25:22	U	SMPL
J4581-14 - 1	EP1-Q3-FD-01	725	0.1356	-7471B	PPB		24 Aug 2018 17:27:38	U	SMPL
CCV88 - 1	CCV88	22593	5.0480	-7471B	PPB		24 Aug 2018 17:29:54	U	SMPL
CCB88 - 1	CCB88	-127	-0.0558	-7471B	PPB		24 Aug 2018 17:32:10	U	SMPL
J4581-16 - 1	EP1-Q3-C4	1795	0.3760	-7471B	PPB		24 Aug 2018 17:34:26	U	SMPL
J4581-17 - 1	EP1-Q3-D1	2312	0.4921	-7471B	PPB		24 Aug 2018 17:36:43	U	SMPL
J4581-18 - 1	EP1-Q3-D2	5265	1.1555	-7471B	PPB		24 Aug 2018 17:38:59	U	SMPL
J4581-19 - 1	EP1-Q3-D3	1011	0.1999	-7471B	PPB		24 Aug 2018 17:41:15	U	SMPL
J4581-20 - 1	EP1-Q3-D4	1366	0.2796	-7471B	PPB		24 Aug 2018 17:43:32	U	SMPL
PB112342BL - 1	PBS	12	-0.0245	-7471B	PPB		24 Aug 2018 17:45:49	U	SMPL
PB112342BS - 1	LCSS	16308	3.6362	-7471B	PPB		24 Aug 2018 17:48:06	U	SMPL
J4270-01 - 1	BU-03-082318	1194	0.2410	-7471B	PPB		24 Aug 2018 17:50:24	U	SMPL
J4270-03 - 1	BU-01-082318	2303	0.4901	-7471B	PPB		24 Aug 2018 17:52:42	U	SMPL
J4273-01 - 1	NB-07-082318	1388	0.2846	-7471B	PPB		24 Aug 2018 17:54:58	U	SMPL
CCV89 - 1	CCV89	22592	5.0478	-7471B	PPB		24 Aug 2018 17:57:14	U	SMPL
CCB89 - 1	CCB89	-133	-0.0571	-7471B	PPB		24 Aug 2018 17:59:31	U	SMPL
J4282-11 - 1	SU-02-082318-A	2887	0.6213	-7471B	PPB		24 Aug 2018 18:01:48	U	SMPL
J4282-13 - 1	SU-02-082318-B	2917	0.6280	-7471B	PPB		24 Aug 2018 18:04:05	U	SMPL
J4621-02 - 1	SP-1	2159	0.4578	-7471B	PPB		24 Aug 2018 18:06:21	U	SMPL
J4621-06 - 1	SP-2	1465	0.3019	-7471B	PPB		24 Aug 2018 18:08:38	U	SMPL
J4629-01 - 1	FK-PS-01-038	641	0.1168	-7471B	PPB		24 Aug 2018 18:10:54	U	SMPL
J4636-01 - 1	BYR-2-E(0-1)	8908	1.9738	-7471B	PPB		24 Aug 2018 18:13:11	U	SMPL
J4636-02 - 1	BYR-2-E(6-7)	1487	0.3068	-7471B	PPB		24 Aug 2018 18:15:29	U	SMPL
J4636-03 - 1	BYR-2-E(9-10)	1347	0.2754	-7471B	PPB		24 Aug 2018 18:17:46	U	SMPL
J4636-03DUP - 1	BYR-2-E(9-10)DUP	1310	0.2670	-7471B	PPB		24 Aug 2018 18:20:04	U	SMPL
J4636-03MS - 1	BYR-2-E(9-10)MS	18428	4.1124	-7471B	PPB		24 Aug 2018 18:22:20	U	SMPL
CCV90 - 1	CCV90	22684	5.0684	-7471B	PPB		24 Aug 2018 18:24:37	U	SMPL
CCB90 - 1	CCB90	-152	-0.0614	-7471B	PPB		24 Aug 2018 18:26:53	U	SMPL
J4636-03MSD - 1	BYR-2-E(9-10)MSD	18599	4.1508	-7471B	PPB		24 Aug 2018 18:29:10	U	SMPL
J4581-10LX5 - 1		183	0.0139	-7471B	PPB		24 Aug 2018 18:31:26	U	SMPL
J4581-10A - 1		20448	4.5662	-7471B	PPB		24 Aug 2018 18:33:43	U	SMPL
J4636-03LX5 - 1		122	0.0002	-7471B	PPB		24 Aug 2018 18:36:00	U	SMPL
J4636-03A - 1		19956	4.4556	-7471B	PPB		24 Aug 2018 18:38:16	U	SMPL
CCV91 - 1	CCV91	22405	5.0058	-7471B	PPB		24 Aug 2018 18:40:33	U	SMPL
CCB91 - 1	CCB91	-184	-0.0686	-7471B	PPB		24 Aug 2018 18:42:50	U	SMPL