



PERCENT SOLID

Supervisor: sweta
Analyst: JIGNESH
Date: 8/27/2018

OVENTEMP IN Celsius(°C): 108
Time IN: 17:15
In Date: 08/24/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M Oven-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:27
Out Date: 08/25/2018
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M Sc-1

QC:LB97547

Lab ID	Client Sample ID	Dish#	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish + Dry Sample Wt(g) (C)	% Solid
J4280-01	PL-05-082418-A	1	1.13	9.52	8.45	87.2
J4280-03	PL-05-082418-B	2	1.16	9.79	8.57	85.9
J4646-01	EP1-Q4-A5	3	1.13	9.81	9.26	93.7
J4646-02	EP1-Q4-A6	4	1.15	9.8	9.4	95.4
J4646-03	EP1-Q4-A7	5	1.18	9.84	9.31	93.9
J4646-04	J4646-03MS	6	1.18	9.84	9.31	93.9
J4646-05	J4646-03MSD	7	1.18	9.84	9.31	93.9
J4646-06	EP1-Q4-B5	8	1.15	9.8	9.2	93.1
J4646-07	EP1-Q4-B6	9	1.16	9.54	9.16	95.5
J4646-08	EP1-Q4-B7	10	1.1	9.66	9.19	94.5
J4646-09	EP1-Q4-C5	11	1.19	9.98	9.45	94.0
J4646-10	EP1-Q4-C6	12	1.13	9.57	9.16	95.1
J4646-11	EP1-Q4-C7	13	1.15	9.97	9.41	93.7
J4646-12	EP1-Q4-D5	14	1.12	9.72	9.22	94.2
J4646-13	EP1-Q4-D6	15	1.17	9.94	9.54	95.4
J4646-14	EP1-Q4-D7	16	1.16	9.6	8.88	91.5
J4646-15	EP1-Q4-FD01	17	1.14	9.74	9.31	95.0
J4648-01	SB-1-Q-3-4-FT	18	1.12	9.97	3.16	23.1
J4648-02	SB-1-Q-8-FT	19	1.11	9.66	7.6	75.9
J4648-03	SB-2-Q-5-FT	20	1.13	9.98	8.15	79.3
J4648-04	SB-2-Q-8-FT	21	1.1	9.71	7.47	74.0
J4648-05	SB-3-Q-5-FT	22	1.14	9.98	5.84	53.2
J4648-06	SB-3-Q-8-FT	23	1.13	9.79	7.41	72.5
J4648-07	SB-4-Q-5-6-FT	24	1.14	9.63	7.88	79.4
J4648-08	SB-4-Q-8-FT	25	1.18	9.96	8.36	81.8
J4652-01	SONO-S	26	1.15	9.95	7.05	67.0
J4652-02	SONO-S	27	1.15	9.95	7.05	67.0
J4655-01	A360952	28	1.00	2.00	2.00	100.0
J4655-02	B370984	29	1.00	2.00	2.00	100.0

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$