


7470A

LB97552
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35378



A= 0.0000e+000

B= 2.7977e-004

C= 8.7202e-003

Rho= 0.9997608

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	-4	0.000	-4				
0.2	0.200	0.207	0.007	707	0.0 %	707				
2.5	2.500	2.464	-0.036	8776	0.0 %	8776				
5.0	5.000	4.952	-0.048	17670	0.0 %	17670				
7.5	7.500	7.663	0.163	27360	0.0 %	27360				
10.0	10.000	9.906	-0.094	35378	0.0 %	35378				

LB97552 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Aug 2018 09:18:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-4	-	0.0000	7470A	PPB	27 Aug 2018 09:25:54	S	Std
0.2 - 1	50.2	707	-	0.2000	7470A	PPB	27 Aug 2018 09:28:02	S	Std
2.5 - 1	52.5	8776	-	2.5000	7470A	PPB	27 Aug 2018 09:30:11	S	Std
5.0 - 1	55	17670	-	5.0000	7470A	PPB	27 Aug 2018 09:32:19	S	Std
7.5 - 1	57.5	27360	-	7.5000	7470A	PPB	27 Aug 2018 09:34:27	S	Std
10.0 - 1	510	35378	-	10.0000	7470A	PPB	27 Aug 2018 09:36:35	S	Std
ICV31 - 1	ICV31	14030	3.9339	-	7470A	PPB	27 Aug 2018 09:39:16	U	SMPL
ICB31 - 1	ICB31	-174	-0.0400	-	7470A	PPB	27 Aug 2018 09:41:22	U	SMPL
CCV92 - 1	CCV92	18031	5.0533	-	7470A	PPB	27 Aug 2018 09:43:30	U	SMPL
CCB92 - 1	CCB92	-174	-0.0400	-	7470A	PPB	27 Aug 2018 09:45:36	U	SMPL
CRA - 1	CRA	663	0.1942	-	7470A	PPB	27 Aug 2018 09:47:44	U	SMPL
HighStd - 1	HighStd	35912	10.0558	-	7470A	PPB	27 Aug 2018 09:49:51	U	SMPL
ChkStd - 1	ChkStd	25956	7.2704	-	7470A	PPB	27 Aug 2018 09:51:58	U	SMPL
PB112364BL - 1	PBW	3	0.0096	-	7470A	PPB	27 Aug 2018 09:59:45	U	SMPL
PB112364BS - 1	LCSW	13747	3.8547	-	7470A	PPB	27 Aug 2018 10:01:52	U	SMPL
PB112364TB - 1	PB112338TB	254	0.0798	-	7470A	PPB	27 Aug 2018 10:08:39	U	SMPL
J4270-02 - 1	BU-03-082318	288	0.0893	-	7470A	PPB	27 Aug 2018 10:10:46	U	SMPL
J4270-02DUP - 1	BU-03-082318DUP	307	0.0946	-	7470A	PPB	27 Aug 2018 10:12:55	U	SMPL
J4270-02MS - 1	BU-03-082318MS	13551	3.7999	-	7470A	PPB	27 Aug 2018 10:15:03	U	SMPL
J4270-02MSD - 1	BU-03-082318MSD	14192	3.9792	-	7470A	PPB	27 Aug 2018 10:17:10	U	SMPL
CCV93 - 1	CCV93	18059	5.0611	-	7470A	PPB	27 Aug 2018 10:19:18	U	SMPL
CCB93 - 1	CCB93	-170	-0.0388	-	7470A	PPB	27 Aug 2018 10:21:25	U	SMPL
J4270-04 - 1	BU-01-082318	222	0.0708	-	7470A	PPB	27 Aug 2018 10:23:33	U	SMPL
J4273-02 - 1	NB-07-082318	222	0.0708	-	7470A	PPB	27 Aug 2018 10:28:29	U	SMPL
J4282-12 - 1	SU-02-082318-A	70	0.0283	-	7470A	PPB	27 Aug 2018 10:30:36	U	SMPL
J4282-14 - 1	SU-02-082318-B	173	0.0571	-	7470A	PPB	27 Aug 2018 10:32:43	U	SMPL
J4620-01 - 1	DRUM-1	69	0.0280	-	7470A	PPB	27 Aug 2018 10:34:50	U	SMPL
J4620-02 - 1	DRUM-2	286	0.0887	-	7470A	PPB	27 Aug 2018 10:36:58	U	SMPL
J4620-03 - 1	DRUM-3	112	0.0401	-	7470A	PPB	27 Aug 2018 10:39:06	U	SMPL
J4621-05 - 1	SP-1	122	0.0429	-	7470A	PPB	27 Aug 2018 10:41:14	U	SMPL
J4621-09 - 1	SP-2	264	0.0826	-	7470A	PPB	27 Aug 2018 10:43:22	U	SMPL
PB112365BL - 1	PBW	41	0.0202	-	7470A	PPB	27 Aug 2018 10:45:29	U	SMPL
CCV94 - 1	CCV94	18050	5.0586	-	7470A	PPB	27 Aug 2018 10:47:36	U	SMPL
CCB94 - 1	CCB94	-158	-0.0355	-	7470A	PPB	27 Aug 2018 10:49:43	U	SMPL
PB112365BS - 1	LCSW	13290	3.7269	-	7470A	PPB	27 Aug 2018 10:51:51	U	SMPL
J4616-01 - 1	VB27184	3471	0.9798	-	7470A	PPB	27 Aug 2018 10:53:58	U	SMPL
J4616-01DUP - 1	VB27184DUP	3003	0.8489	-	7470A	PPB	27 Aug 2018 10:56:05	U	SMPL
J4616-01MS - 1	VB27184MS	18134	5.0821	-	7470A	PPB	27 Aug 2018 10:58:13	U	SMPL
J4616-01MSD - 1	VB27184MSD	18293	5.1266	-	7470A	PPB	27 Aug 2018 11:00:21	U	SMPL
J4617-01 - 1	N48969	-137	-0.0296	-	7470A	PPB	27 Aug 2018 11:02:30	U	SMPL
J4617-04 - 1	N1332	32	0.0177	-	7470A	PPB	27 Aug 2018 11:04:40	U	SMPL
J4618-01 - 1	FM1026	218	0.0697	-	7470A	PPB	27 Aug 2018 11:06:48	U	SMPL
J4621-10 - 1	OILY-WATER	9476	2.6598	-	7470A	PPB	27 Aug 2018 11:08:57	U	SMPL
J4627-01 - 1	VB-N29937	366	0.1111	-	7470A	PPB	27 Aug 2018 11:11:04	U	SMPL
CCV95 - 1	CCV95	17836	4.9987	-	7470A	PPB	27 Aug 2018 11:13:13	U	SMPL
CCB95 - 1	CCB95	-62	-0.0086	-	7470A	PPB	27 Aug 2018 11:15:20	U	SMPL
J4638-01 - 1	PS-N48964	80	0.0311	-	7470A	PPB	27 Aug 2018 11:17:27	U	SMPL
J4642-01 - 1	2727	127	0.0443	-	7470A	PPB	27 Aug 2018 11:19:35	U	SMPL
J4652-04 - 1	SONO-W	109833	30.7367	-	7470A	PPB	27 Aug 2018 11:21:42	U	SMPL
J4270-02LX5 - 1		35	0.0185	-	7470A	PPB	27 Aug 2018 11:26:24	U	SMPL
J4270-02A - 1		15030	4.2137	-	7470A	PPB	27 Aug 2018 11:28:31	U	SMPL
J4616-01LX5 - 1		447	0.1338	-	7470A	PPB	27 Aug 2018 11:30:39	U	SMPL
J4616-01A - 1		20592	5.7697	-	7470A	PPB	27 Aug 2018 11:32:47	U	SMPL
J4652-04DLX5 - 1		19972	5.5963	-	7470A	PPB	27 Aug 2018 11:34:55	U	SMPL
CCV96 - 1	CCV96	17863	5.0063	-	7470A	PPB	27 Aug 2018 11:37:05	U	SMPL
CCB96 - 1	CCB96	-83	-0.0145	-	7470A	PPB	27 Aug 2018 11:39:14	U	SMPL