

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 08/27/2018

TEMP1 IN: 103 08/27/2018 07:00 TEMP OUT: 104 08/27/2018 08:15

TEMP2 IN: 104 08/27/2018 09:00 TEMP OUT: 104 08/27/2018 10:10

TEMP3 IN: 105 08/27/2018 12:15 TEMP OUT: 105 08/27/2018 14:00

TEMP4 IN: 105 08/27/2018 14:40 TEMP OUT: 105 08/27/2018 16:15

Run Number: LB97554

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB97554BL	LB97554BL	1.4015	1.4015	100	1.4015	1.4015	1.4015	0.0000	0
2	LB97554BS	LB97554BS	1.4000	1.4000	100	1.4500	1.4500	1.4500	0.0500	500
3	J4659-01	082218-TIPC-F-D-B	1.3975	1.3975	200	1.4005	1.4005	1.4005	0.0030	15
4	J4659-01DUP	082218-TIPC-F-D-BDUP	1.3970	1.3970	200	1.4001	1.4001	1.4001	0.0031	15.5
5	J4659-02	082218-TIPC-F-D-S	1.3884	1.3884	300	1.3940	1.3940	1.3940	0.0056	18.7
6	J4659-03	082218-TIPC-F-U-B	1.3884	1.3884	300	1.3932	1.3932	1.3932	0.0048	16
7	J4659-04	082218-TIPC-F-U-S	1.3750	1.3750	400	1.3792	1.3792	1.3792	0.0042	10.5
8	J4659-05	082218-SIDI-E-D-B	1.4035	1.4035	400	1.4086	1.4086	1.4086	0.0051	12.8
9	J4659-06	082218-SIDI-E-D-S	1.3946	1.3946	400	1.3983	1.3983	1.3983	0.0037	9.3
10	J4659-07	082218-SIDI-E-U-B	1.4044	1.4044	400	1.4130	1.4130	1.4130	0.0086	21.5
11	J4659-08	082218-SIDI-E-U-S	1.3914	1.3914	400	1.3964	1.3964	1.3964	0.0050	12.5
12	J4659-09	082218-DBRI-E-D-B	1.3899	1.3899	400	1.4030	1.4030	1.4030	0.0131	32.8
13	J4659-10	082218-DBRI-E-D-M	1.4007	1.4007	400	1.4064	1.4064	1.4064	0.0057	14.3
14	J4659-11	082218-DBRI-E-D-S	1.3879	1.3879	400	1.3919	1.3919	1.3919	0.0040	10
15	J4659-12	082218-DBRI-E-U-B	1.4055	1.4055	400	1.4121	1.4121	1.4121	0.0066	16.5



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 08/27/2018TEMP1 IN: 103 08/27/2018 07:00 TEMP OUT: 104 08/27/2018 08:15TEMP2 IN: 104 08/27/2018 09:00 TEMP OUT: 104 08/27/2018 10:10TEMP3 IN: 105 08/27/2018 12:15 TEMP OUT: 105 08/27/2018 14:00TEMP4 IN: 105 08/27/2018 14:40 TEMP OUT: 105 08/27/2018 16:15Run Number: LB97554BalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	J4659-13	082218-DBRI-E-U-M	1.4067	1.4067	400	1.4114	1.4114	1.4114	0.0047	11.8
17	J4659-14	082218-DBRI-E-U-S	1.4007	1.4007	200	1.4040	1.4040	1.4040	0.0033	16.5
18	J4659-14DUP	082218-DBRI-E-U-SDUP	1.3990	1.3990	200	1.4022	1.4022	1.4022	0.0032	16

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS082718-2

WorkList ID : 116071

Date : 8/27/2018 9:19:49 AM

21397554

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/28/2018	Water	J4659-01 C	TSS	Cool 4 deg C	HDR103	E51	082218-TIPC-F-D-B	08/22/2018	SM2540 D
08/28/2018	Water	J4659-02 C	TSS	Cool 4 deg C	HDR103	E51	082218-TIPC-F-D-S	08/22/2018	SM2540 D
08/28/2018	Water	J4659-03 C	TSS	Cool 4 deg C	HDR103	E51	082218-TIPC-F-U-B	08/22/2018	SM2540 D
08/28/2018	Water	J4659-04 C	TSS	Cool 4 deg C	HDR103	E51	082218-TIPC-F-U-S	08/22/2018	SM2540 D
08/29/2018	Water	J4659-05 C	TSS	Cool 4 deg C	HDR103	E51	082218-SIDI-E-D-B	08/22/2018	SM2540 D
08/29/2018	Water	J4659-06 C	TSS	Cool 4 deg C	HDR103	E51	082218-SIDI-E-D-S	08/22/2018	SM2540 D
08/29/2018	Water	J4659-07 C	TSS	Cool 4 deg C	HDR103	E51	082218-SIDI-E-U-B	08/22/2018	SM2540 D
08/29/2018	Water	J4659-08 C	TSS	Cool 4 deg C	HDR103	E51	082218-SIDI-E-U-S	08/22/2018	SM2540 D
08/29/2018	Water	J4659-09 C	TSS	Cool 4 deg C	HDR103	E51	082218-DBRI-E-D-B	08/22/2018	SM2540 D
08/29/2018	Water	J4659-10 C	TSS	Cool 4 deg C	HDR103	E51	082218-DBRI-E-D-M	08/22/2018	SM2540 D
08/29/2018	Water	J4659-11 C	TSS	Cool 4 deg C	HDR103	E51	082218-DBRI-E-D-S	08/22/2018	SM2540 D
08/29/2018	Water	J4659-12 C	TSS	Cool 4 deg C	HDR103	E51	082218-DBRI-E-U-B	08/22/2018	SM2540 D
08/29/2018	Water	J4659-13 C	TSS	Cool 4 deg C	HDR103	E51	082218-DBRI-E-U-M	08/22/2018	SM2540 D
08/29/2018	Water	J4659-14 C	TSS	Cool 4 deg C	HDR103	E51	082218-DBRI-E-U-S	08/22/2018	SM2540 D

FP-7642602

Date/Time 08/27/18, 08:00
 Received by: AS
 Relinquished by: NP

Date/Time 08/27/18, 12:00
 Received by: NP
 Relinquished by: AS