



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB97559

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.0	80.6	80.6	08/27/2018	10:25
2	J4618-02		1	7	>100.0	>212.0	>212.0	08/27/2018	10:55
3	J4618-02DUP		1	7	>100.0	>212.0	>212.0	08/27/2018	11:20
4	J4652-05		1	8	>100.0	>212.0	>212.0	08/27/2018	11:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 97559

WorkList Name : FP82718

WorkList ID : 116079

Date : 8/27/2018 9:46:17 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/28/2018	Water	J4618-02	Flash Point	Cool 4 deg C	PSEG01	K63	FM1026	08/23/2018	1010A
08/29/2018	Water	J4652-05	Flash Point	Cool 4 deg C	PSEG01	E22	SONO-W	08/24/2018	1010A

Date/Time 08/27/18 9:50 AM
 Received by: RS
 Relinquished by: CP

Date/Time 08/27/18
 Received by:
 Relinquished by: