


7471B

LB97563
INSTRUMENT ID: CV1

Linear

 μ Abs.:

47375



A= 0.0000e+000

B= 2.1248e-004

C= -9.8871e-005

Rho= 0.9999330

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	39	0.000	39				
0.2	0.200	0.232	0.032	1090	0.0 %	1090				
2.5	2.500	2.496	-0.004	11747	0.0 %	11747				
5.0	5.000	4.962	-0.038	23354	0.0 %	23354				
7.5	7.500	7.436	-0.064	34998	0.0 %	34998				
10.0	10.000	10.066	0.066	47375	0.0 %	47375				

LB97563 INSTRUMENT ID :: CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Aug 2018 14:17:04

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7471B	PPB	27 Aug 2018 14:24:13	S	Std
0.2 - 1	50.2	1090	-	0.2000	7471B	PPB	27 Aug 2018 14:26:31	S	Std
2.5 - 1	52.5	11747	-	2.5000	7471B	PPB	27 Aug 2018 14:28:49	S	Std
5.0 - 1	55	23354	-	5.0000	7471B	PPB	27 Aug 2018 14:31:06	S	Std
7.5 - 1	57.5	34998	-	7.5000	7471B	PPB	27 Aug 2018 14:33:23	S	Std
10.0 - 1	510	47375	-	10.0000	7471B	PPB	27 Aug 2018 14:38:35	S	Std
ICV32 - 1	ICV32	18928	4.0217	-	7471B	PPB	27 Aug 2018 14:42:38	U	SMPL
ICB32 - 1	ICB32	33	0.0069	-	7471B	PPB	27 Aug 2018 14:44:54	U	SMPL
CCV01 - 1	CCV01	23897	5.0775	-	7471B	PPB	27 Aug 2018 14:47:10	U	SMPL
CCB01 - 1	CCB01	1	0.0001	-	7471B	PPB	27 Aug 2018 14:49:26	U	SMPL
CRA - 1	CRA	1059	0.2249	-	7471B	PPB	27 Aug 2018 14:51:43	U	SMPL
HighStd - 1	HighStd	47753	10.1464	-	7471B	PPB	27 Aug 2018 14:53:58	U	SMPL
ChkStd - 1	ChkStd	32797	6.9686	-	7471B	PPB	27 Aug 2018 14:56:14	U	SMPL
PB112379BL - 1	PBS	63	0.0133	-	7471B	PPB	27 Aug 2018 14:58:33	U	SMPL
PB112379BS - 1	LCSS	18453	3.9208	-	7471B	PPB	27 Aug 2018 15:00:51	U	SMPL
J4280-01 - 1	PL-05-082418-A	10624	2.2573	-	7471B	PPB	27 Aug 2018 15:03:08	U	SMPL
J4280-03 - 1	PL-05-082418-B	12436	2.6423	-	7471B	PPB	27 Aug 2018 15:05:26	U	SMPL
J4646-01 - 1	EP1-Q4-A5	1732	0.3679	-	7471B	PPB	27 Aug 2018 15:07:44	U	SMPL
J4646-02 - 1	EP1-Q4-A6	436	0.0925	-	7471B	PPB	27 Aug 2018 15:10:02	U	SMPL
J4646-03 - 1	EP1-Q4-A7	670	0.1423	-	7471B	PPB	27 Aug 2018 15:12:18	U	SMPL
CCV02 - 1	CCV02	23147	4.9181	-	7471B	PPB	27 Aug 2018 15:14:34	U	SMPL
CCB02 - 1	CCB02	30	0.0063	-	7471B	PPB	27 Aug 2018 15:16:50	U	SMPL
J4646-03DUP - 1	EP1-Q4-A7DUP	817	0.1735	-	7471B	PPB	27 Aug 2018 15:26:25	U	SMPL
J4646-04 - 1	J4646-03MS	17402	3.6975	-	7471B	PPB	27 Aug 2018 15:28:40	U	SMPL
J4646-05 - 1	J4646-03MSD	18043	3.8337	-	7471B	PPB	27 Aug 2018 15:30:56	U	SMPL
J4646-06 - 1	EP1-Q4-B5	1426	0.3029	-	7471B	PPB	27 Aug 2018 15:33:12	U	SMPL
J4646-07 - 1	EP1-Q4-B6	477	0.1013	-	7471B	PPB	27 Aug 2018 15:35:30	U	SMPL
J4646-08 - 1	EP1-Q4-B7	1026	0.2179	-	7471B	PPB	27 Aug 2018 15:37:47	U	SMPL
J4646-09 - 1	EP1-Q4-C5	1530	0.3250	-	7471B	PPB	27 Aug 2018 15:40:04	U	SMPL
J4646-10 - 1	EP1-Q4-C6	885	0.1879	-	7471B	PPB	27 Aug 2018 15:42:21	U	SMPL
J4646-11 - 1	EP1-Q4-C7	1598	0.3394	-	7471B	PPB	27 Aug 2018 15:44:39	U	SMPL
J4646-12 - 1	EP1-Q4-D5	1020	0.2166	-	7471B	PPB	27 Aug 2018 15:46:55	U	SMPL
CCV03 - 1	CCV03	23414	4.9749	-	7471B	PPB	27 Aug 2018 15:49:11	U	SMPL
CCB03 - 1	CCB03	28	0.0059	-	7471B	PPB	27 Aug 2018 15:51:27	U	SMPL
J4646-13 - 1	EP1-Q4-D6	1079	0.2292	-	7471B	PPB	27 Aug 2018 15:53:44	U	SMPL
J4646-14 - 1	EP1-Q4-D7	589	0.1251	-	7471B	PPB	27 Aug 2018 15:56:00	U	SMPL
J4646-15 - 1	EP1-Q4-FD01	806	0.1712	-	7471B	PPB	27 Aug 2018 15:58:16	U	SMPL
J4652-01 - 1	SONO-S	10966	2.3299	-	7471B	PPB	27 Aug 2018 16:00:32	U	SMPL
J4646-03LX5 - 1		228	0.0483	-	7471B	PPB	27 Aug 2018 16:05:48	U	SMPL
J4646-03A - 1		18634	3.9592	-	7471B	PPB	27 Aug 2018 16:08:05	U	SMPL
CCV04 - 1	CCV04	23249	4.9398	-	7471B	PPB	27 Aug 2018 16:10:22	U	SMPL
CCB04 - 1	CCB04	41	0.0086	-	7471B	PPB	27 Aug 2018 16:12:39	U	SMPL