

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/27/2018 16:22

Test: Total CN ug/d

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	98.714	0.0	0.094	
ICB1	-0.039	0.0	0.004	
CCV1	246.871	0.0	0.229	
CCB1	1.137	0.0	0.005	
PB112402BL	0.720	0.0	0.005	
PB112402BS	100.529	0.0	0.095	
LOWPB112402	10.876	0.0	0.014	
HIGHPB112402	487.044	0.0	0.447	
J4280-01	2.513	0.0	0.006	
J4280-01DUP	1.623	0.0	0.005	
J4280-01MS	35.108	0.0	0.036	
J4280-01MSD	34.953	0.0	0.036	
J4280-03	2.175	0.0	0.006	
J4652-01	29.354	0.0	0.031	
CCV2	246.644	0.0	0.229	
CCB2	0.838	0.0	0.005	
J4661-02	1314.170	0.0	1.201	Test limit high
PB112403BL	0.811	0.0	0.005	
PB112403BS	100.404	0.0	0.095	
J4618-01	16.829	0.0	0.019	
J4618-01DUP	16.806	0.0	0.019	
J4618-01MS	51.109	0.0	0.050	
J4618-01MSD	51.599	0.0	0.051	
J4652-04	35.915	0.0	0.037	
CCV3	252.944	0.0	0.234	
CCB3	-0.228	0.0	0.004	
J4661-02DLX5	256.921	0.0	0.238	
CCV4	268.085	0.0	0.248	
CCB4	2.752	0.0	0.006	

N 29
Mean 126.454
SD 258.0616
CV% 204.07

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092
Reviewed by : AS

8/27/2018 14:07


Test Total CN

Accepted 8/27/2018 14:07

Factor 1098
Bias 0.004

Coeff. of det. 0.999948

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN	0.003	-1.0200	0.0000	
2	5.0PPBCN	0.007	3.8493	5.0000	
3	10PPBCN	0.013	9.4823	10.0000	
4	50PPBCN	0.050	50.4599	50.0000	
5	100PPBCN	0.096	101.6302	100.0000	
6	250PPBCN	0.233	251.8966	250.0000	
7	500PPBCN	0.458	498.7014	500.0000	

AS

Aquakem v. 7.2AQ1

Results from time period:

Mon Aug 27 15:05:11 2018

Mon Aug 27 16:21:49 2018

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPBCN	A	Total CN	P	-1.02	µg/l	8/27/2018 14:06:12	
5.0PPBCN	A	Total CN	P	3.8493	µg/l	8/27/2018 14:06:13	
10PPBCN	A	Total CN	P	9.4823	µg/l	8/27/2018 14:06:14	
50PPBCN	A	Total CN	P	50.4599	µg/l	8/27/2018 14:06:15	
100PPBCN	A	Total CN	P	101.6302	µg/l	8/27/2018 14:06:16	
250PPBCN	A	Total CN	P	251.8966	µg/l	8/27/2018 14:06:17	
500PPBCN	A	Total CN	P	498.7014	µg/l	8/27/2018 14:06:18	
ICV1	S	Total CN	P	98.7136	µg/l	8/27/2018 15:05:11	
ICB1	S	Total CN	P	-0.0394	µg/l	8/27/2018 15:05:12	
CCV1	S	Total CN	P	246.8713	µg/l	8/27/2018 15:05:13	
CCB1	S	Total CN	P	1.1366	µg/l	8/27/2018 15:05:14	
PB112402BL	S	Total CN	P	0.7201	µg/l	8/27/2018 15:05:15	
PB112402BS	S	Total CN	P	100.5288	µg/l	8/27/2018 15:05:16	
LOWPB112402	S	Total CN	P	10.8763	µg/l	8/27/2018 15:05:17	
HIGHPB112402	S	Total CN	P	487.0439	µg/l	8/27/2018 15:05:18	
J4280-01	S	Total CN	P	2.513	µg/l	8/27/2018 15:05:19	
J4280-01DUP	S	Total CN	P	1.6234	µg/l	8/27/2018 15:05:20	
J4280-01MS	S	Total CN	P	35.1084	µg/l	8/27/2018 15:05:21	
J4280-01MSD	S	Total CN	P	34.9526	µg/l	8/27/2018 15:12:43	
J4280-03	S	Total CN	P	2.1755	µg/l	8/27/2018 15:12:44	
J4652-01	S	Total CN	P	29.3536	µg/l	8/27/2018 15:12:45	
CCV2	S	Total CN	P	246.6443	µg/l	8/27/2018 15:40:36	
CCB2	S	Total CN	P	0.8382	µg/l	8/27/2018 15:40:37	
J4661-02	S	Total CN	P	1314.17	µg/l	8/27/2018 15:40:38	
PB112403BL	S	Total CN	P	0.8106	µg/l	8/27/2018 15:40:39	
PB112403BS	S	Total CN	P	100.4035	µg/l	8/27/2018 15:40:40	
J4618-01	S	Total CN	P	16.8286	µg/l	8/27/2018 15:40:41	
J4618-01DUP	S	Total CN	P	16.806	µg/l	8/27/2018 15:40:42	
J4618-01MS	S	Total CN	P	51.1094	µg/l	8/27/2018 15:40:43	
J4618-01MSD	S	Total CN	P	51.5994	µg/l	8/27/2018 15:40:44	
J4652-04	S	Total CN	P	35.9148	µg/l	8/27/2018 15:40:45	
CCV3	S	Total CN	P	252.9439	µg/l	8/27/2018 15:40:46	
CCB3	S	Total CN	P	-0.2282	µg/l	8/27/2018 15:44:01	
J4661-02DLX5	S	Total CN	P	256.9211	µg/l	8/27/2018 16:21:47	
CCV4	S	Total CN	P	268.0852	µg/l	8/27/2018 16:21:48	
CCB4	S	Total CN	P	2.7523	µg/l	8/27/2018 16:21:49	