



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : apatel

Run Number: LB97573

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	08/27/2018	16:40
2	CAL2	1	Water			21.9	7.01	08/27/2018	16:42
3	CAL3	1	Water			21.8	10.04	08/27/2018	16:44
4	ICV	1	Water			21.7	7.00	08/27/2018	16:46
5	CCV1	1	Water			21.6	2.00	08/27/2018	16:48
6	J4664-01	1	Solid	20.02	20	23.3	7.04	08/27/2018	16:50
7	J4664-01DUP	1	Solid	20.03	20	23.4	7.05	08/27/2018	16:52
8	CCV2	1	Water			22.0	12.00	08/27/2018	17:08

LB 97573

WORKLIST(Hardcopy Internal Chain)

WorkList Name : pH082718

WorkList ID : 116105

Date : 8/27/2018 4:01:39 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/25/2018	Solid	J4664-01	pH	Cool 4 deg C	CTIA01	H22	WJ-18-08-02-TSHG-02	08/25/2018	9045D

Date/Time 08/27/18 4:30
Received by: SP
Relinquished by: CP

Date/Time 08/27/18 5:15
Received by: CP
Relinquished by: SP