



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97574

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	08/27/2018	16:40
2	CAL2	1	Water			21.9	7.01	08/27/2018	16:42
3	CAL3	1	Water			21.8	10.04	08/27/2018	16:44
4	ICV	1	Water			21.7	7.00	08/27/2018	16:46
5	CCV1	1	Water			21.6	2.00	08/27/2018	16:48
6	J4274-06	1	Solid	20.02	20	24.5	8.45	08/27/2018	16:54
7	J4274-06DUP	1	Solid	20.04	20	24.6	8.46	08/27/2018	16:56
8	J4276-06	1	Solid	20.03	20	24.4	8.88	08/27/2018	16:58
9	CCV2	1	Water			22.0	12.00	08/27/2018	17:08

LBAT SAH

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity082718-1

WorkList ID : 116095

Date : 8/27/2018 3:17:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/27/2018	Solid	J4274-06	Corrosivity	Cool 4 deg C	PSEG05		HD-01-082718	08/27/2018	9040C
08/27/2018	Solid	J4276-06	Corrosivity	Cool 4 deg C	PSEG05		OK-02-082718	08/27/2018	9040C

Date/Time 08/27/18 4:30
Received by: SP
Relinquished by: CP

Date/Time 08/27/18 5:25
Received by: CP
Relinquished by: SP