



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB97575

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	08/27/2018	16:40
2	CAL2	1	Water			21.9	7.01	08/27/2018	16:42
3	CAL3	1	Water			21.8	10.04	08/27/2018	16:44
4	ICV	1	Water			21.7	7.00	08/27/2018	16:46
5	CCV1	1	Water			21.6	2.00	08/27/2018	16:48
6	J4277-10	1	Solid	20.02	20	24.2	8.36	08/27/2018	17:00
7	J4277-10DUP	1	Solid	20.03	20	24.3	8.37	08/27/2018	17:02
8	J4277-12	1	Solid	20.02	20	24.9	8.00	08/27/2018	17:04
9	J4671-02	1	Solid	20.02	20	23.7	6.34	08/27/2018	17:06
10	CCV2	1	Water			22.0	12.00	08/27/2018	17:08

97575
B

WORKLIST(Hardcopy Internal Chain)

WorkList Name : corrosivity082718 WorkList ID : 116094 Date : 8/27/2018 3:17:29 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/27/2018	Solid	J4277-10	Corrosivity	Cool 4 deg C	PSEG05		OR-01-082718-A	08/27/2018	9045D
08/27/2018	Solid	J4277-12	Corrosivity	Cool 4 deg C	PSEG05		OR-01-082718-B	08/27/2018	9045D
08/27/2018	Solid	J4671-02	Corrosivity	Cool 4 deg C	PSEG01	K63	TL-SPEEDY-DRY	08/27/2018	9045D

Date/Time 08/27/18 4:30
 Received by: SP
 Relinquished by: CP

Date/Time 08/27/18 5:15
 Received by: CP
 Relinquished by: VSP