


7471B

LB97584
INSTRUMENT ID: CV1

Linear

 μ Abs.:

43889



A= 0.0000e+000

B= 2.2678e-004

C= -8.8925e-003

Rho= 0.9999399

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	38	0.000	38				
0.2	0.200	0.181	-0.019	836	0.0 %	836				
2.5	2.500	2.508	0.008	11099	0.0 %	11099				
5.0	5.000	4.989	-0.011	22039	0.0 %	22039				
7.5	7.500	7.578	0.078	33457	0.0 %	33457				
10.0	10.000	9.944	-0.056	43889	0.0 %	43889				

LB97584 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Aug 2018 10:39:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	38	-	0.0000	7471B	PPB	28 Aug 2018 10:45:53	S	Std
0.2 - 1	50.2	836	-	0.2000	7471B	PPB	28 Aug 2018 10:48:10	S	Std
2.5 - 1	52.5	11099	-	2.5000	7471B	PPB	28 Aug 2018 10:50:28	S	Std
5.0 - 1	55	22039	-	5.0000	7471B	PPB	28 Aug 2018 10:52:45	S	Std
7.5 - 1	57.5	33457	-	7.5000	7471B	PPB	28 Aug 2018 10:55:03	S	Std
10.0 - 1	510	43889	-	10.0000	7471B	PPB	28 Aug 2018 10:59:57	S	Std
ICV34 - 1	ICV34	17400	3.9370	-	7471B	PPB	28 Aug 2018 11:02:38	U	SMPL
ICB34 - 1	ICB34	-126	-0.0375	-	7471B	PPB	28 Aug 2018 11:04:55	U	SMPL
CCV10 - 1	CCV10	22205	5.0267	-	7471B	PPB	28 Aug 2018 11:07:12	U	SMPL
CCB10 - 1	CCB10	-111	-0.0341	-	7471B	PPB	28 Aug 2018 11:09:28	U	SMPL
CRA - 1	CRA	865	0.1873	-	7471B	PPB	28 Aug 2018 11:11:46	U	SMPL
HighStd - 1	HighStd	44409	10.0620	-	7471B	PPB	28 Aug 2018 11:14:02	U	SMPL
ChkStd - 1	ChkStd	31638	7.1659	-	7471B	PPB	28 Aug 2018 11:16:18	U	SMPL
PB112426BL - 1	PBS	-206	-0.0556	-	7471B	PPB	28 Aug 2018 11:18:36	U	SMPL
PB112426BS - 1	LCSS	15792	3.5724	-	7471B	PPB	28 Aug 2018 11:20:54	U	SMPL
J4274-05 - 1	HD-01-082718	2999	0.6712	-	7471B	PPB	28 Aug 2018 11:28:06	U	SMPL
J4274-05DUP - 1	HD-01-082718DUP	2859	0.6395	-	7471B	PPB	28 Aug 2018 11:30:23	U	SMPL
J4274-05MS - 1	HD-01-082718MS	18599	4.2089	-	7471B	PPB	28 Aug 2018 11:32:41	U	SMPL
J4274-05MSD - 1	HD-01-082718MSD	19545	4.4235	-	7471B	PPB	28 Aug 2018 11:34:59	U	SMPL
J4276-05 - 1	OK-02-082718	1458	0.3217	-	7471B	PPB	28 Aug 2018 11:37:14	U	SMPL
CCV11 - 1	CCV11	21714	4.9153	-	7471B	PPB	28 Aug 2018 11:39:30	U	SMPL
CCB11 - 1	CCB11	-117	-0.0354	-	7471B	PPB	28 Aug 2018 11:41:46	U	SMPL
J4277-09 - 1	OR-01-082718-A	2791	0.6240	-	7471B	PPB	28 Aug 2018 11:44:02	U	SMPL
J4277-11 - 1	OR-01-082718-B	1116	0.2442	-	7471B	PPB	28 Aug 2018 11:46:18	U	SMPL
J4661-02 - 1	WASTE-CLASS-HOT-SPOT	48110	10.9013	-	7471B	PPB	28 Aug 2018 11:48:34	U	SMPL
J4664-01 - 1	WJ-18-08-02-TSHG-02	1189	0.2607	-	7471B	PPB	28 Aug 2018 11:50:50	U	SMPL
J4671-01 - 1	TL-SPEEDY-DRY	339	0.0680	-	7471B	PPB	28 Aug 2018 11:53:09	U	SMPL
J4274-05LX5 - 1		566	0.1195	-	7471B	PPB	28 Aug 2018 11:55:27	U	SMPL
J4274-05A - 1		21415	4.8475	-	7471B	PPB	28 Aug 2018 11:57:44	U	SMPL
J4661-02DLX2 - 1		23997	5.4331	-	7471B	PPB	28 Aug 2018 12:00:02	U	SMPL
CCV12 - 1	CCV12	21991	4.9781	-	7471B	PPB	28 Aug 2018 12:02:22	U	SMPL
CCB12 - 1	CCB12	-136	-0.0397	-	7471B	PPB	28 Aug 2018 12:04:39	U	SMPL