


245.1

LB 97594
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35284



A= 0.0000e+000

B= 2.8410e-004

C= 1.8437e-002

Rho= 0.9998903

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	-126	0.000	-126				
0.05	0.050	0.048	-0.002	103	0.0 %	103				
0.2	0.200	0.222	0.022	718	0.0 %	718				
2.5	2.500	2.482	-0.018	8671	0.0 %	8671				
5.0	5.000	5.081	0.081	17819	0.0 %	17819				
7.5	7.500	7.392	-0.108	25954	0.0 %	25954				
10.0	10.000	10.043	0.043	35284	0.0 %	35284				

LB97594 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Aug 2018 13:51:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-126	-	0.0000	245.1	PPB	28 Aug 2018 14:19:42	S	Std
0.05 - 1	50.05	103	-	0.0500	245.1	PPB	28 Aug 2018 14:22:00	S	Std
0.2 - 1	50.2	718	-	0.2000	245.1	PPB	28 Aug 2018 14:24:18	S	Std
2.5 - 1	52.5	8671	-	2.5000	245.1	PPB	28 Aug 2018 14:26:35	S	Std
5.0 - 1	55	17819	-	5.0000	245.1	PPB	28 Aug 2018 14:28:53	S	Std
7.5 - 1	57.5	25954	-	7.5000	245.1	PPB	28 Aug 2018 14:31:09	S	Std
10.0 - 1	510	35284	-	10.0000	245.1	PPB	28 Aug 2018 14:33:27	S	Std
ICV35 - 1	ICV35	13550	3.8680	-245.1		PPB	28 Aug 2018 14:36:25	U	SMPL
ICB35 - 1	ICB35	-93	-0.0080	-245.1		PPB	28 Aug 2018 14:38:41	U	SMPL
CCV13 - 1	CCV13	17871	5.0956	-245.1		PPB	28 Aug 2018 14:40:58	U	SMPL
CCB13 - 1	CCB13	-84	-0.0054	-245.1		PPB	28 Aug 2018 14:43:14	U	SMPL
CRA - 1	CRA	156	0.0628	-245.1		PPB	28 Aug 2018 14:48:04	U	SMPL
HighStd - 1	HighStd	35213	10.0224	-245.1		PPB	28 Aug 2018 14:53:08	U	SMPL
ChkStd - 1	ChkStd	23768	6.7709	-245.1		PPB	28 Aug 2018 14:55:24	U	SMPL
PB112417BL - 1	PBW	-18	0.0133	-245.1		PPB	28 Aug 2018 15:00:23	U	SMPL
PB112417BS - 1	LCSW	13549	3.8677	-245.1		PPB	28 Aug 2018 15:02:39	U	SMPL
J4651-01 - 1	JH-BTHRM-SINK	323	0.1102	-245.1		PPB	28 Aug 2018 15:04:56	U	SMPL
J4651-02 - 1	JH-SHWR	365	0.1221	-245.1		PPB	28 Aug 2018 15:07:13	U	SMPL
J4651-03 - 1	NR-BTHRM-SINK	396	0.1309	-245.1		PPB	28 Aug 2018 15:09:31	U	SMPL
J4651-04 - 1	NR-SHWR	495	0.1591	-245.1		PPB	28 Aug 2018 15:11:49	U	SMPL
J4651-05 - 1	NR-KIT-SINK-DW	746	0.2304	-245.1		PPB	28 Aug 2018 15:14:04	U	SMPL
CCV14 - 1	CCV14	17936	5.1140	-245.1		PPB	28 Aug 2018 15:16:20	U	SMPL
CCB14 - 1	CCB14	-171	-0.0301	-245.1		PPB	28 Aug 2018 15:18:36	U	SMPL
J4656-01 - 1	082418-SIDI-F-D-B	-5	0.0170	-245.1		PPB	28 Aug 2018 15:20:53	U	SMPL
J4656-02 - 1	082418-SIDI-F-D-M	42	0.0304	-245.1		PPB	28 Aug 2018 15:23:09	U	SMPL
J4656-03 - 1	082418-SIDI-F-D-S	35	0.0284	-245.1		PPB	28 Aug 2018 15:25:25	U	SMPL
J4656-04 - 1	082418-SIDI-F-U-B	88	0.0434	-245.1		PPB	28 Aug 2018 15:27:41	U	SMPL
J4656-05 - 1	082418-SIDI-F-U-M	54	0.0338	-245.1		PPB	28 Aug 2018 15:29:58	U	SMPL
J4656-06 - 1	082418-SIDI-F-U-S	40	0.0298	-245.1		PPB	28 Aug 2018 15:32:15	U	SMPL
J4656-07 - 1	082418-DBRI-F-D-B	48	0.0321	-245.1		PPB	28 Aug 2018 15:34:32	U	SMPL
J4656-08 - 1	082418-DBRI-F-D-M	94	0.0451	-245.1		PPB	28 Aug 2018 15:36:49	U	SMPL
J4656-09 - 1	082418-DBRI-F-D-S	59	0.0352	-245.1		PPB	28 Aug 2018 15:39:08	U	SMPL
CCV15 - 1	CCV15	17777	5.0689	-245.1		PPB	28 Aug 2018 15:43:40	U	SMPL
CCB15 - 1	CCB15	-123	-0.0165	-245.1		PPB	28 Aug 2018 15:45:57	U	SMPL
J4656-11 - 1	082418-DBRI-F-U-M	-52	0.0037	-245.1		PPB	28 Aug 2018 15:48:14	U	SMPL
J4656-12 - 1	082418-DBRI-F-U-S	-10	0.0156	-245.1		PPB	28 Aug 2018 15:50:30	U	SMPL
J4656-12DUP - 1	082418-DBRI-F-U-SDUP	-2	0.0179	-245.1		PPB	28 Aug 2018 15:52:46	U	SMPL
J4656-12MS - 1	082418-DBRI-F-U-SMS	15140	4.3197	-245.1		PPB	28 Aug 2018 15:55:02	U	SMPL
J4656-12MSD - 1	082418-DBRI-F-U-SMSD	13451	3.8399	-245.1		PPB	28 Aug 2018 15:59:56	U	SMPL
PB112418BL - 1	PBW	-36	0.0082	-245.1		PPB	28 Aug 2018 16:04:56	U	SMPL
PB112418BS - 1	LCSW	13711	3.9137	-245.1		PPB	28 Aug 2018 16:07:12	U	SMPL
J4659-01 - 1	082218-TIPC-F-D-B	-166	-0.0287	-245.1		PPB	28 Aug 2018 16:09:30	U	SMPL
J4659-02 - 1	082218-TIPC-F-D-S	-29	0.0102	-245.1		PPB	28 Aug 2018 16:11:48	U	SMPL
J4659-03 - 1	082218-TIPC-F-U-B	9	0.0210	-245.1		PPB	28 Aug 2018 16:14:05	U	SMPL
CCV16 - 1	CCV16	17800	5.0754	-245.1		PPB	28 Aug 2018 16:16:21	U	SMPL
CCB16 - 1	CCB16	-167	-0.0290	-245.1		PPB	28 Aug 2018 16:18:38	U	SMPL
J4659-04 - 1	082218-TIPC-F-U-S	-34	0.0088	-245.1		PPB	28 Aug 2018 16:20:55	U	SMPL
J4659-05 - 1	082218-SIDI-E-D-B	8	0.0207	-245.1		PPB	28 Aug 2018 16:23:12	U	SMPL
J4659-06 - 1	082218-SIDI-E-D-S	29	0.0267	-245.1		PPB	28 Aug 2018 16:25:28	U	SMPL
J4659-07 - 1	082218-SIDI-E-U-B	-16	0.0139	-245.1		PPB	28 Aug 2018 16:27:45	U	SMPL
J4659-08 - 1	082218-SIDI-E-U-S	-15	0.0142	-245.1		PPB	28 Aug 2018 16:30:02	U	SMPL
J4659-09 - 1	082218-DBRI-E-D-B	21	0.0244	-245.1		PPB	28 Aug 2018 16:32:19	U	SMPL
J4659-10 - 1	082218-DBRI-E-D-M	20	0.0241	-245.1		PPB	28 Aug 2018 16:34:36	U	SMPL
J4659-11 - 1	082218-DBRI-E-D-S	-7	0.0164	-245.1		PPB	28 Aug 2018 16:36:53	U	SMPL
J4659-12 - 1	082218-DBRI-E-U-B	44	0.0309	-245.1		PPB	28 Aug 2018 16:39:11	U	SMPL
J4659-13 - 1	082218-DBRI-E-U-M	26	0.0258	-245.1		PPB	28 Aug 2018 16:41:27	U	SMPL
CCV17 - 1	CCV17	17409	4.9643	-245.1		PPB	28 Aug 2018 16:43:44	U	SMPL
CCB17 - 1	CCB17	-151	-0.0245	-245.1		PPB	28 Aug 2018 16:46:00	U	SMPL
J4659-14 - 1	082218-DBRI-E-U-S	0	0.0184	-245.1		PPB	28 Aug 2018 16:48:18	U	SMPL
J4659-14DUP - 1	082218-DBRI-E-U-SDUP	22	0.0247	-245.1		PPB	28 Aug 2018 16:50:35	U	SMPL
J4659-14MS - 1	082218-DBRI-E-U-SMS	14961	4.2689	-245.1		PPB	28 Aug 2018 16:52:51	U	SMPL
J4659-14MSD - 1	082218-DBRI-E-U-SMSD	15238	4.3475	-245.1		PPB	28 Aug 2018 16:55:07	U	SMPL
PB112419BL - 1	PBW	-109	-0.0125	-245.1		PPB	28 Aug 2018 16:57:24	U	SMPL
PB112419BS - 1	LCSW	13164	3.7583	-245.1		PPB	28 Aug 2018 16:59:43	U	SMPL
J4660-01 - 1	082318-SIDI-F-D-B	-110	-0.0128	-245.1		PPB	28 Aug 2018 17:02:00	U	SMPL
J4660-02 - 1	082318-SIDI-F-D-M	29	0.0267	-245.1		PPB	28 Aug 2018 17:04:18	U	SMPL
J4660-03 - 1	082318-SIDI-F-D-S	7	0.0204	-245.1		PPB	28 Aug 2018 17:06:36	U	SMPL
J4660-04 - 1	082318-SIDI-F-U-B	53	0.0335	-245.1		PPB	28 Aug 2018 17:08:52	U	SMPL
CCV18 - 1	CCV18	17581	5.0132	-245.1		PPB	28 Aug 2018 17:11:09	U	SMPL
CCB18 - 1	CCB18	-161	-0.0273	-245.1		PPB	28 Aug 2018 17:13:25	U	SMPL
J4660-05 - 1	082318-SIDI-F-U-M	-7	0.0164	-245.1		PPB	28 Aug 2018 17:15:42	U	SMPL
J4660-06 - 1	082318-DBRI-F-U-S	46	0.0315	-245.1		PPB	28 Aug 2018 17:17:59	U	SMPL
J4660-07 - 1	082318-DBRI-F-D-B	41	0.0301	-245.1		PPB	28 Aug 2018 17:20:15	U	SMPL
J4660-08 - 1	082318-DBRI-F-D-M	9	0.0210	-245.1		PPB	28 Aug 2018 17:22:31	U	SMPL
J4660-09 - 1	082318-DBRI-F-D-S	16	0.0230	-245.1		PPB	28 Aug 2018 17:24:48	U	SMPL
J4660-10 - 1	082318-DBRI-F-U-B	7	0.0204	-245.1		PPB	28 Aug 2018 17:27:05	U	SMPL
J4660-11 - 1	082318-DBRI-F-U-M	28	0.0264	-245.1		PPB	28 Aug 2018 17:29:22	U	SMPL
J4660-12 - 1	082318-DBRI-F-U-S	5	0.0199	-245.1		PPB	28 Aug 2018 17:31:40	U	SMPL
J4660-13 - 1	082318-TIPC-F-D-B	93	0.0449	-245.1		PPB	28 Aug 2018 17:33:58	U	SMPL
J4660-14 - 1	082318-TIPC-F-D-S	15	0.0227	-245.1		PPB	28 Aug 2018 17:36:15	U	SMPL
CCV19 - 1	CCV19	17619	5.0240	-245.1		PPB	28 Aug 2018 17:38:32	U	SMPL

LB97594 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 28 Aug 2018 13:51:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
CCB19 - 1	CCB19	-148	-0.0236		-245.1	PPB	28 Aug 2018 17:40:49	U	SMPL
J4660-15 - 1	082318-TIPC-F-U-B	-27	0.0108		-245.1	PPB	28 Aug 2018 17:43:07	U	SMPL
J4660-16 - 1	082318-TIPC-F-U-S	-20	0.0128		-245.1	PPB	28 Aug 2018 17:51:57	U	SMPL
J4660-16DUP - 1	082318-TIPC-F-U-SDUP	66	0.0372		-245.1	PPB	28 Aug 2018 17:54:14	U	SMPL
J4660-16MS - 1	082318-TIPC-F-U-SMS	15664	4.4686		-245.1	PPB	28 Aug 2018 17:56:31	U	SMPL
J4660-16MSD - 1	082318-TIPC-F-U-SMSD	15270	4.3566		-245.1	PPB	28 Aug 2018 17:58:48	U	SMPL
CCV20 - 1	CCV20	17755	5.0626		-245.1	PPB	28 Aug 2018 18:01:05	U	SMPL
CCB20 - 1	CCB20	-216	-0.0429		-245.1	PPB	28 Aug 2018 18:03:22	U	SMPL
J4656-12LX5 - 1		47	0.0318		-245.1	PPB	28 Aug 2018 18:05:41	U	SMPL
J4659-14LX5 - 1		-38	0.0076		-245.1	PPB	28 Aug 2018 18:10:15	U	SMPL
J4659-14A - 1		13705	3.9120		-245.1	PPB	28 Aug 2018 18:12:32	U	SMPL
J4660-16LX5 - 1		-34	0.0088		-245.1	PPB	28 Aug 2018 18:14:49	U	SMPL
J4660-16A - 1		15944	4.5481		-245.1	PPB	28 Aug 2018 18:17:08	U	SMPL
J4656-12A - 1		13518	3.8589		-245.1	PPB	28 Aug 2018 18:22:36	U	SMPL
J4656-10 - 1	082418-DBRIF-U-B	-84	-0.0054		-245.1	PPB	28 Aug 2018 18:24:52	U	SMPL
CCV21 - 1	CCV21	17345	4.9461		-245.1	PPB	28 Aug 2018 18:27:12	U	SMPL
CCB21 - 1	CCB21	-184	-0.0338		-245.1	PPB	28 Aug 2018 18:29:29	U	SMPL