



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : AMANDEEP

Run Number: LB97598

Slope : 98.8

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.7	4.01	08/28/2018	14:20
2	CAL2	1	Water			21.8	7.01	08/28/2018	14:22
3	CAL3	1	Water			22.0	10.04	08/28/2018	14:24
4	ICV	1	Water			21.9	7.00	08/28/2018	14:26
5	CCV1	1	Water			21.9	2.00	08/28/2018	14:28
6	J4663-02	1	Solid	20.02	20	24.0	9.01	08/28/2018	14:30
7	J4677-01	1	Solid	20.02	20	24.6	8.15	08/28/2018	14:32
8	J4677-02	1	Solid	20.03	20	24.3	8.44	08/28/2018	14:34
9	J4677-03	1	Solid	20.02	20	24.3	7.66	08/28/2018	14:36
10	J4677-03DUP	1	Solid	20.03	20	24.4	7.67	08/28/2018	14:38
11	CCV2	1	Water			22.0	12.00	08/28/2018	14:40

WORKLIST(Hardcopy Internal Chain)

UB97598

WorkList Name : corrosivity082818

WorkList ID : 116149

Date : 8/28/2018 2:06:46 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/28/2018	Solid	J4663-02	Corrosivity	Cool 4 deg C	PSEG01	K63	6722AA	08/28/2018	9045D
08/27/2018	Solid	J4677-01	Corrosivity	Cool 4 deg C	USAR03	G42	CB-0818-S-TCLP-GRID-42(1)	08/27/2018	9045D
08/27/2018	Solid	J4677-02	Corrosivity	Cool 4 deg C	USAR03	G42	CB-0818-S-TCLP-GRID-42(2)	08/27/2018	9045D
08/27/2018	Solid	J4677-03	Corrosivity	Cool 4 deg C	USAR03	G42	CB-0818-S-TCLP-GRID-42(3)	08/27/2018	9045D

Date/Time 08/28/18 2:10

Received by: SP

Relinquished by: CP

Date/Time 08/28/18 3:00

Received by: CP

Relinquished by: SP