

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AparnaParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: AMANDEEPRun Number: LB97615

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP65054
calibration std. phosphate 1 ppm	WP67390
calibration std. phosphate 0.5 ppm	WP67391
calibration std. phosphate 0.3 ppm	WP67392
calibration std. 0 ppm	WP67395
phosphate ICV std.	WP67399
phosphate CCV std.	WP67398
calibration std. phosphate 0.1 ppm	WP67393
calibration std. phosphate 0.05 ppm	WP67394
Combined reagent	WP67397
11 N sulfuric acid	WP66905
Phenolphthalein indicator	WP66907
Sodium hydroxide, 1N	WP66924

Intercept: -0.0011Slope: 0.6496Regression: 0.999849

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	08/29/2018	13:05
2	CAL2	0.05	1	50	50	0.032	08/29/2018	13:05
3	CAL3	0.10	1	50	50	0.065	08/29/2018	13:06
4	CAL4	0.30	1	50	50	0.186	08/29/2018	13:07
5	CAL5	0.50	1	50	50	0.329	08/29/2018	13:07
6	CAL6	1.00	1	50	50	0.648	08/29/2018	13:07



Analytical Summary Report

Analysis Method: 365.3

ANALYST: Aparna

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: AMANDEEP

Run Number: LB97615

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.496	08/29/2018	13:08
2	ICB		1	50	50	0.000	0.002	08/29/2018	13:08
3	CCV1	0.50	1	50	50	0.327	0.505	08/29/2018	13:09
4	CCB1		1	50	50	0.000	0.002	08/29/2018	13:09
5	PB112473BL		1	50	50	0.000	0.002	08/29/2018	13:11
6	PB112473BS	0.50	1	50	50	0.323	0.499	08/29/2018	13:11
7	J4640-01		1	50	50	0.384	0.593	08/29/2018	13:12
8	J4640-01DUP		1	50	50	0.396	0.611	08/29/2018	13:12
9	J4640-01MS	0.50	1	50	50	0.641	0.988	08/29/2018	13:13
10	J4640-01MSD	0.50	1	50	50	0.644	0.993	08/29/2018	13:13
11	CCV2	0.50	1	50	50	0.326	0.504	08/29/2018	13:14
12	CCB2		1	50	50	0.000	0.002	08/29/2018	13:14