



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB97622

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	08/29/2018	12:20
2	CAL2	1	Water			21.1	7.00	08/29/2018	12:22
3	CAL3	1	Water			21.9	10.04	08/29/2018	12:24
4	ICV	1	Water			22.0	7.01	08/29/2018	12:26
5	CCV1	1	Water			21.8	2.00	08/29/2018	12:28
6	J4690-01	1	Solid	20.02	20	24.2	7.40	08/29/2018	12:30
7	J4690-01DUP	1	Solid	20.03	20	24.3	7.41	08/29/2018	12:32
8	CCV2	1	Water			22.0	12.00	08/29/2018	12:34

LBQ-622

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH082918

WorkList ID : 116183

Date : 8/29/2018 12:08:27 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/28/2018	Solid	J4690-01	pH	Cool 4 deg C	CTIA01	H32	WJ-18-08-28-TSHG-03	08/28/2018	9045D

Date/Time 08/29/18 12:10
Received by: SP
Relinquished by: CP

Date/Time 08/29/18 12:45
Received by: CP
Relinquished by: SP