



Analytical Summary Report

Analysis Method: 9040C

Analyst By : Sweta

Parameter: Corrosivity

Supervisor Review By : AMANDEEP

Run Number: LB97624

Slope : 98.9

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.1	4.01	08/29/2018	16:20
2	CAL2	1	Water			22.0	7.00	08/29/2018	16:22
3	CAL3	1	Water			21.9	10.04	08/29/2018	16:24
4	ICV	1	Water			22.0	7.01	08/29/2018	16:26
5	CCV1	1	Water			21.8	2.00	08/29/2018	16:28
6	J4275-20	1	Solid	20.02	20	26.0	7.24	08/29/2018	16:30
7	J4275-20DUP	1	Solid	20.03	20	26.1	7.25	08/29/2018	16:32
8	J4275-22	1	Solid	20.02	20	25.8	7.13	08/29/2018	16:34
9	CCV2	1	Water			21.9	12.00	08/29/2018	16:48

WORKLIST(Hardcopy Internal Chain)

LB97624

WorkList Name : corrosivity082918

WorkList ID : 116195

Date : 8/29/2018 2:31:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/29/2018	Solid	J4275-20	Corrosivity	Cool 4 deg C	PSEG05		CL-02-082918-A	08/29/2018	9040C
08/29/2018	Solid	J4275-22	Corrosivity	Cool 4 deg C	PSEG05		CL-02-082918-B	08/29/2018	9040C

Date/Time 08/29/18 4:00

Received by: SP

Relinquished by: CP

Date/Time 08/29/18 5:00

Received by: CP

Relinquished by: SP