

ISM02.4


LB 97626

INSTRUMENT ID: CV2

Linear

 μ Abs.:

42233



A= 0.0000e+000

B= 2.3888e-004

C= -9.7187e-002

Rho= 0.9999605

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.047	-0.047	209	0.000	209					1
0.2	0.200	0.208	0.008	1277	0.0 %	1277					4
2.5	2.500	2.534	0.034	11015	0.0 %	11015					1
5.0	5.000	5.045	0.045	21525	0.0 %	21525					1
7.5	7.500	7.469	-0.031	31676	0.0 %	31676					0
10.0	10.000	9.991	-0.009	42233	0.0 %	42233					0

ms
8129118

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	209	-		0 ISM02.4	PPB	8/29/2018 17:20	Std
	0.2 S0.2	1277	-		0.2 ISM02.4	PPB	8/29/2018 17:22	Std
	2.5 S2.5	11015	-		2.5 ISM02.4	PPB	8/29/2018 17:25	Std
	5 S5	21525	-		5 ISM02.4	PPB	8/29/2018 17:29	Std
	7.5 S7.5	31676	-		7.5 ISM02.4	PPB	8/29/2018 17:32	Std
	10 S10.0	42233	-		10 ISM02.4	PPB	8/29/2018 17:36	Std
ICV	ICV	17181	4.0069 -		ISM02.4	PPB	8/29/2018 17:39	SMPL
ICB	ICB	56	-0.0838 -		ISM02.4	PPB	8/29/2018 17:41	SMPL
CCV032	CCV032	21249	4.9787 -		ISM02.4	PPB	8/29/2018 17:43	SMPL
CCB032	CCB032	48	-0.0857 -		ISM02.4	PPB	8/29/2018 17:45	SMPL
PB112475BL	PBW003	182	-0.0537 -		ISM02.4	PPB	8/29/2018 17:48	SMPL
J4626-01	MDB3H4	17155	4.0007 -		ISM02.4	PPB	8/29/2018 17:50	SMPL
J4626-19	MDB3G4	1	-0.0969 -		ISM02.4	PPB	8/29/2018 17:52	SMPL
J4626-20	MDB3G5	109	-0.0711 -		ISM02.4	PPB	8/29/2018 17:54	SMPL
J4626-21	MDB3G6	110	-0.0709 -		ISM02.4	PPB	8/29/2018 17:57	SMPL
J4626-22	MDB3G6D	161	-0.0587 -		ISM02.4	PPB	8/29/2018 17:59	SMPL
J4626-23	MDB3G6S	4448	0.9653 -		ISM02.4	PPB	8/29/2018 18:01	SMPL
CCV033	CCV033	21206	4.9684 -		ISM02.4	PPB	8/29/2018 18:08	SMPL
CCB033	CCB033	32	-0.0895 -		ISM02.4	PPB	8/29/2018 18:10	SMPL