


7470A

LB97628
INSTRUMENT ID: CV1 μ Abs.:

38088



Linear

A= 0.0000e+000

B= 2.6569e-004

C= -1.0563e-002

Rho= 0.9997964

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.028	0.028	147	0.000	147				
0.2	0.200	0.205	0.005	812	0.0 %	812				
2.5	2.500	2.542	0.042	9606	0.0 %	9606				
5.0	5.000	4.938	-0.062	18627	0.0 %	18627				
7.5	7.500	7.377	-0.123	27807	0.0 %	27807				
10.0	10.000	10.109	0.109	38088	0.0 %	38088				

LB97628 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 30 Aug 2018 08:57:45

Sample ID	Extended ID	μ Abs	Conc	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	147	-	0.00007470A	PPB	30 Aug 2018 09:01:46	S	Std
0.2 - 1	50.2	812	-	0.20007470A	PPB	30 Aug 2018 09:03:55	S	Std
2.5 - 1	52.5	9606	-	2.50007470A	PPB	30 Aug 2018 09:06:03	S	Std
5.0 - 1	55	18627	-	5.00007470A	PPB	30 Aug 2018 09:08:10	S	Std
7.5 - 1	57.5	27807	-	7.50007470A	PPB	30 Aug 2018 09:10:18	S	Std
10.0 - 1	510	38088	-	10.00007470A	PPB	30 Aug 2018 09:12:26	S	Std
ICV38 - 1	ICV38	14775	3.9150	-7470A	PPB	30 Aug 2018 09:15:02	U	SMPL
ICB38 - 1	ICB38	-18	-0.0153	-7470A	PPB	30 Aug 2018 09:17:08	U	SMPL
CCV34 - 1	CCV34	18927	5.0181	-7470A	PPB	30 Aug 2018 09:19:16	U	SMPL
CCB34 - 1	CCB34	0	-0.0106	-7470A	PPB	30 Aug 2018 09:21:23	U	SMPL
CRA - 1	CRA	838	0.2121	-7470A	PPB	30 Aug 2018 09:23:31	U	SMPL
HighStd - 1	HighStd	37089	9.8435	-7470A	PPB	30 Aug 2018 09:25:38	U	SMPL
ChkStd - 1	ChkStd	25849	6.8572	-7470A	PPB	30 Aug 2018 09:27:45	U	SMPL
PB112482BL - 1	PBW	-105	-0.0385	-7470A	PPB	30 Aug 2018 09:29:55	U	SMPL
PB112482BS - 1	LCSW	13803	3.6567	-7470A	PPB	30 Aug 2018 09:32:03	U	SMPL
PB112482TB - 1	PB112451TB	2	-0.0100	-7470A	PPB	30 Aug 2018 09:34:11	U	SMPL
J4663-02 - 1	6722AA	166	0.0335	-7470A	PPB	30 Aug 2018 09:36:19	U	SMPL
J4663-02DUP - 1	6722AADUP	168	0.0341	-7470A	PPB	30 Aug 2018 09:38:27	U	SMPL
J4663-02MS - 1	6722AAMS	12589	3.3342	-7470A	PPB	30 Aug 2018 09:40:36	U	SMPL
J4663-02MSD - 1	6722AAMSD	13378	3.5438	-7470A	PPB	30 Aug 2018 09:42:43	U	SMPL
CCV35 - 1	CCV35	19046	5.0497	-7470A	PPB	30 Aug 2018 09:44:50	U	SMPL
CCB35 - 1	CCB35	-1	-0.0108	-7470A	PPB	30 Aug 2018 09:46:58	U	SMPL
J4677-01 - 1	CB-0818-S-TCLP-GRID-42(1)	179	0.0370	-7470A	PPB	30 Aug 2018 09:49:06	U	SMPL
J4677-02 - 1	CB-0818-S-TCLP-GRID-42(2)	165	0.0333	-7470A	PPB	30 Aug 2018 09:51:13	U	SMPL
J4677-03 - 1	CB-0818-S-TCLP-GRID-42(3)	177	0.0365	-7470A	PPB	30 Aug 2018 09:53:20	U	SMPL
J4663-02LX5 - 1		129	0.0237	-7470A	PPB	30 Aug 2018 09:55:27	U	SMPL
J4663-02A - 1		15568	4.1257	-7470A	PPB	30 Aug 2018 09:57:35	U	SMPL
CCV36 - 1	CCV36	18557	4.9198	-7470A	PPB	30 Aug 2018 09:59:42	U	SMPL
CCB36 - 1	CCB36	-96	-0.0361	-7470A	PPB	30 Aug 2018 10:01:51	U	SMPL