


245.1

LB97631
INSTRUMENT ID: CV1

Linear

 μ Abs.:

34480



A= 0.0000e+000

B= 2.8583e-004

C= -1.7826e-002

Rho= 0.9996330

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.017	-0.017	2	0.000	2				
0.05	0.050	0.048	-0.002	232	0.0 %	232				
0.2	0.200	0.197	-0.003	751	0.0 %	751				
2.5	2.500	2.447	-0.053	8624	0.0 %	8624				
5.0	5.000	5.037	0.037	17685	0.0 %	17685				
7.5	7.500	7.700	0.200	27001	0.0 %	27001				
10.0	10.000	9.838	-0.162	34480	0.0 %	34480				

LB97631 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 30 Aug 2018 12:02:38

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	2	-	0.0000	245.1	PPB	30 Aug 2018 12:05:21	S	Std
0.05 - 1	SO.05	232	-	0.0500	245.1	PPB	30 Aug 2018 12:07:39	S	Std
0.2 - 1	SO.2	751	-	0.2000	245.1	PPB	30 Aug 2018 12:12:38	S	Std
2.5 - 1	SO.25	8624	-	2.5000	245.1	PPB	30 Aug 2018 12:14:56	S	Std
5.0 - 1	SO.5	17685	-	5.0000	245.1	PPB	30 Aug 2018 12:17:13	S	Std
7.5 - 1	SO.75	27001	-	7.5000	245.1	PPB	30 Aug 2018 12:19:30	S	Std
10.0 - 1	SO.10	34480	-	10.0000	245.1	PPB	30 Aug 2018 12:24:34	S	Std
ICV39 - 1	ICV39	13506	3.8426	-245.1		PPB	30 Aug 2018 12:27:23	U	SMPL
ICB39 - 1	ICB39	-105	-0.0478	-245.1		PPB	30 Aug 2018 12:29:38	U	SMPL
CCV37 - 1	CCV37	17717	5.0463	-245.1		PPB	30 Aug 2018 12:31:55	U	SMPL
CCB37 - 1	CCB37	-143	-0.0587	-245.1		PPB	30 Aug 2018 12:34:11	U	SMPL
CRA - 1	CRA	197	0.0385	-245.1		PPB	30 Aug 2018 12:36:28	U	SMPL
HighStd - 1	HighStd	34355	9.8019	-245.1		PPB	30 Aug 2018 12:38:44	U	SMPL
ChkStd - 1	ChkStd	24115	6.8750	-245.1		PPB	30 Aug 2018 12:41:00	U	SMPL
PB112483BL - 1	PBW	-36	-0.0281	-245.1		PPB	30 Aug 2018 12:46:01	U	SMPL
PB112483BS - 1	LCSW	13742	3.9101	-245.1		PPB	30 Aug 2018 12:48:18	U	SMPL
J4683-01 - 1	082718-JETT-F-D-B	43	-0.0055	-245.1		PPB	30 Aug 2018 12:50:35	U	SMPL
J4683-02 - 1	082718-JETT-F-D-S	147	0.0242	-245.1		PPB	30 Aug 2018 12:52:53	U	SMPL
J4683-03 - 1	082718-JETT-F-U-B	190	0.0365	-245.1		PPB	30 Aug 2018 12:55:11	U	SMPL
J4683-04 - 1	082718-JETT-F-U-S	189	0.0362	-245.1		PPB	30 Aug 2018 12:57:29	U	SMPL
J4683-05 - 1	082718-SIDI-F-D-B	200	0.0393	-245.1		PPB	30 Aug 2018 12:59:45	U	SMPL
CCV38 - 1	CCV38	17897	5.0977	-245.1		PPB	30 Aug 2018 13:02:01	U	SMPL
CCB38 - 1	CCB38	-129	-0.0547	-245.1		PPB	30 Aug 2018 13:04:17	U	SMPL
J4683-06 - 1	082718-SIDI-F-D-S	170	0.0308	-245.1		PPB	30 Aug 2018 13:06:35	U	SMPL
J4683-07 - 1	082718-SIDI-F-U-B	222	0.0456	-245.1		PPB	30 Aug 2018 13:08:51	U	SMPL
J4683-08 - 1	082718-SIDI-F-U-S	144	0.0233	-245.1		PPB	30 Aug 2018 13:11:07	U	SMPL
J4683-09 - 1	082718-DBRI-F-D-B	224	0.0462	-245.1		PPB	30 Aug 2018 13:13:23	U	SMPL
J4683-10 - 1	082718-DBRI-F-D-M	219	0.0448	-245.1		PPB	30 Aug 2018 13:15:40	U	SMPL
J4683-11 - 1	082718-DBRI-F-D-S	166	0.0296	-245.1		PPB	30 Aug 2018 13:17:56	U	SMPL
J4683-13 - 1	082718-DBRI-F-U-M	192	0.0371	-245.1		PPB	30 Aug 2018 13:22:33	U	SMPL
J4683-14 - 1	082718-DBRI-F-U-S	184	0.0348	-245.1		PPB	30 Aug 2018 13:24:51	U	SMPL
J4683-15 - 1	082718-TIPC-F-D-B	175	0.0322	-245.1		PPB	30 Aug 2018 13:27:07	U	SMPL
CCV39 - 1	CCV39	17550	4.9985	-245.1		PPB	30 Aug 2018 13:29:23	U	SMPL
CCB39 - 1	CCB39	-104	-0.0476	-245.1		PPB	30 Aug 2018 13:31:40	U	SMPL
J4683-16 - 1	082718-TIPC-F-D-S	161	0.0282	-245.1		PPB	30 Aug 2018 13:33:57	U	SMPL
J4683-17 - 1	082718-TIPC-F-U-B	188	0.0359	-245.1		PPB	30 Aug 2018 13:36:14	U	SMPL
J4683-18 - 1	082718-TIPC-F-U-S	188	0.0359	-245.1		PPB	30 Aug 2018 13:38:30	U	SMPL
J4683-18DUP - 1	082718-TIPC-F-U-SDUP	193	0.0373	-245.1		PPB	30 Aug 2018 13:40:47	U	SMPL
J4683-18MS - 1	082718-TIPC-F-U-SMS	15132	4.3074	-245.1		PPB	30 Aug 2018 13:43:03	U	SMPL
J4683-18MSD - 1	082718-TIPC-F-U-SMSD	15177	4.3203	-245.1		PPB	30 Aug 2018 13:45:20	U	SMPL
J4683-18LX5 - 1		37	-0.0072	-245.1		PPB	30 Aug 2018 13:47:37	U	SMPL
J4683-18A - 1		15052	4.2845	-245.1		PPB	30 Aug 2018 13:56:35	U	SMPL
J4683-12 - 1	082718-DBRI-F-U-B	38	-0.0070	-245.1		PPB	30 Aug 2018 13:58:52	U	SMPL
CCV40 - 1	CCV40	17773	5.0623	-245.1		PPB	30 Aug 2018 14:01:09	U	SMPL
CCB40 - 1	CCB40	-118	-0.0516	-245.1		PPB	30 Aug 2018 14:03:25	U	SMPL