



Analytical Summary Report

Analysis Method: 9045D

Analyst By : Sweta

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB97639

Slope : 99.0

BalanceID: WC SC-4

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			21.9	4.01	08/30/2018	13:00
2	CAL2	1	Water			22.0	7.01	08/30/2018	13:02
3	CAL3	1	Water			21.9	10.04	08/30/2018	13:04
4	ICV	1	Water			21.8	7.00	08/30/2018	13:06
5	CCV1	1	Water			22.0	2.00	08/30/2018	13:08
6	J4702-01	1	Solid	20.02	20	23.4	5.60	08/30/2018	13:10
7	J4702-01DUP	1	Solid	20.03	20	23.5	5.61	08/30/2018	13:12
8	J4715-01	1	Solid	20.02	20	21.7	6.85	08/30/2018	13:14
9	CCV2	1	Water			21.8	12.00	08/30/2018	13:16

WORKLIST(Hardcopy Internal Chain)

LB 97639

WorkList Name : PH083018

WorkList ID : 116233

Date : 8/30/2018 10:39:07 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/29/2018	Solid	J4702-01	pH	Cool 4 deg C	TULL02	H52	FK-PS-01-039	08/29/2018	9045D
	Solid	J4715-01	pH	Cool 4 deg C	CTIA01	H43	WJ-18-08-29-TSHG-04	08/29/2018	9045D

Date/Time 08/30/18 12:50
 Received by: SP
 Relinquished by: CP

Date/Time 08/30/18 1:30
 Received by: CP
 Relinquished by: SP