



## Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB97640

| Reagent/Standard       | Lot/Log # |
|------------------------|-----------|
| TURBIDITY CCV 10 NTU   | WP67445   |
| Turbidity std, 40NTU   | WP67440   |
| turbidity std 5 NTU    | WP67443   |
| turbidity std 20 NTU   | WP67441   |
| Turbidity std 1 NTU    | WP67444   |
| turbidity std 0 NTU    | WP67439   |
| turbidity std 10 NTU   | WP67442   |
| 10 NTU Standard 500 ml | W2297     |

Intercept: 0.114

Slope: 0.9784

Regression: 1

| Seq | Lab ID | True Value (NTU) | Dilution | Reading | Anal Date  | Anal Time |
|-----|--------|------------------|----------|---------|------------|-----------|
| 1   | CAL1   | 0                | 1        | 0.231   | 08/30/2018 | 12:06     |
| 2   | CAL2   | 1                | 1        | 0.998   | 08/30/2018 | 12:08     |
| 3   | CAL3   | 5                | 1        | 4.890   | 08/30/2018 | 12:10     |
| 4   | CAL4   | 10               | 1        | 9.920   | 08/30/2018 | 12:12     |
| 5   | CAL5   | 20               | 1        | 19.800  | 08/30/2018 | 12:14     |
| 6   | CAL6   | 40               | 1        | 39.200  | 08/30/2018 | 12:15     |



## Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB97640

| Seq | Lab ID      | True Value<br>(NTU) | Dilution | Reading | Result<br>(NTU) | AnalDate   | Anal Time |
|-----|-------------|---------------------|----------|---------|-----------------|------------|-----------|
| 1   | ICV         | 10                  | 1        | 9.680   | 9.777           | 08/30/2018 | 12:17     |
| 2   | ICB         |                     | 1        | 0.221   | 0.109           | 08/30/2018 | 12:19     |
| 3   | CCV1        | 10                  | 1        | 9.870   | 9.971           | 08/30/2018 | 12:21     |
| 4   | CCB1        |                     | 1        | 0.225   | 0.113           | 08/30/2018 | 12:23     |
| 5   | LB97640BL   |                     | 1        | 0.226   | 0.114           | 08/30/2018 | 12:25     |
| 6   | J4713-01    |                     | 1        | 37.600  | 38.314          | 08/30/2018 | 12:27     |
| 7   | J4713-01DUP |                     | 1        | 37.200  | 37.905          | 08/30/2018 | 12:29     |
| 8   | CCV2        | 10                  | 1        | 9.900   | 10.002          | 08/30/2018 | 12:31     |
| 9   | CCB2        |                     | 1        | 0.216   | 0.104           | 08/30/2018 | 12:33     |

6897640

# WORKLIST(Hardcopy Internal Chain)

WorkList Name : turbidity083018

WorkList ID : 116232

Date : 8/30/2018 10:23:31 AM

| Due Date   | Matrix | Sample   | Test      | Preservative | Customer | Storage Location | Customer Sample | Collect Date | Method   |
|------------|--------|----------|-----------|--------------|----------|------------------|-----------------|--------------|----------|
| 08/31/2018 | Water  | J4713-01 | Turbidity | Cool 4 deg C | LIR001   | H11              | UDS-CH-4-01     | 08/29/2018   | SM2130 B |

Date/Time 8/30/18 11:30  
 Received by: AC  
 Relinquished by: CP

Date/Time 8/30/18 12:45  
 Received by: CP  
 Relinquished by: AC