

7471B


LB97644

INSTRUMENT ID: CY1

Linear

 μ Abs.:

46214



A= 0.0000e+000

B= 2.2041e-004

C= -2.3593e-003

Rho= 0.9992194

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	127	0.000	127				
0.2	0.200	0.207	0.007	952	0.0 %	952				
2.5	2.500	2.553	0.053	11595	0.0 %	11595				
5.0	5.000	5.029	0.029	22829	0.0 %	22829				
7.5	7.500	7.201	-0.299	32682	0.0 %	32682				
10.0	10.000	10.183	0.183	46214	0.0 %	46214				

LB97644 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Aug 2018 16:03:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	127	-	0.0000	7471B	PPB	30 Aug 2018 16:04:14	S	Std
0.2 - 1	50.2	952	-	0.2000	7471B	PPB	30 Aug 2018 16:06:32	S	Std
2.5 - 1	52.5	11595	-	2.5000	7471B	PPB	30 Aug 2018 16:08:50	S	Std
5.0 - 1	55	22829	-	5.0000	7471B	PPB	30 Aug 2018 16:11:07	S	Std
7.5 - 1	57.5	32682	-	7.5000	7471B	PPB	30 Aug 2018 16:13:24	S	Std
10.0 - 1	510	46214	-	10.0000	7471B	PPB	30 Aug 2018 16:15:42	S	Std
ICV41 - 1	ICV41	17891	3.9409	-7471B	PPB	PPB	30 Aug 2018 16:18:29	U	SMPL
ICB41 - 1	ICB41	-32	-0.0094	-7471B	PPB	PPB	30 Aug 2018 16:20:45	U	SMPL
CCV44 - 1	CCV44	23033	5.0742	-7471B	PPB	PPB	30 Aug 2018 16:23:01	U	SMPL
CCB44 - 1	CCB44	-8	-0.0041	-7471B	PPB	PPB	30 Aug 2018 16:25:17	U	SMPL
CRA - 1	CRA	953	0.2077	-7471B	PPB	PPB	30 Aug 2018 16:27:34	U	SMPL
HighStd - 1	HighStd	46503	10.2472	-7471B	PPB	PPB	30 Aug 2018 16:29:50	U	SMPL
ChkStd - 1	ChkStd	30900	6.8082	-7471B	PPB	PPB	30 Aug 2018 16:32:05	U	SMPL
PB112484BL - 1	PBS	-125	-0.0299	-7471B	PPB	PPB	30 Aug 2018 16:34:24	U	SMPL
PB112484BS - 1	LCSS	15785	3.4767	-7471B	PPB	PPB	30 Aug 2018 16:36:42	U	SMPL
J4275-19 - 1	CL-02-082918-A	2060	0.4517	-7471B	PPB	PPB	30 Aug 2018 16:46:26	U	SMPL
J4275-19DUP - 1	CL-02-082918-ADUP	1916	0.4199	-7471B	PPB	PPB	30 Aug 2018 16:48:42	U	SMPL
J4275-19MS - 1	CL-02-082918-AMS	19004	4.1862	-7471B	PPB	PPB	30 Aug 2018 16:51:00	U	SMPL
J4275-19MSD - 1	CL-02-082918-AMSD	19040	4.1942	-7471B	PPB	PPB	30 Aug 2018 16:53:17	U	SMPL
J4275-21 - 1	CL-02-082918-B	2056	0.4508	-7471B	PPB	PPB	30 Aug 2018 16:55:33	U	SMPL
CCV45 - 1	CCV45	22866	5.0374	-7471B	PPB	PPB	30 Aug 2018 16:57:49	U	SMPL
CCB45 - 1	CCB45	-22	-0.0072	-7471B	PPB	PPB	30 Aug 2018 17:00:05	U	SMPL
J4282-15 - 1	SU-01-082918-A	3755	0.8253	-7471B	PPB	PPB	30 Aug 2018 17:02:21	U	SMPL
J4282-17 - 1	SU-01-082918-B	3747	0.8235	-7471B	PPB	PPB	30 Aug 2018 17:04:37	U	SMPL
J4282-19 - 1	SU-01-082918-C	3068	0.6738	-7471B	PPB	PPB	30 Aug 2018 17:06:53	U	SMPL
J4690-01 - 1	WJ-18-08-28-TSHG-03	2112	0.4631	-7471B	PPB	PPB	30 Aug 2018 17:09:10	U	SMPL
J4699-01 - 1	CG-SOIL-1	1100	0.2401	-7471B	PPB	PPB	30 Aug 2018 17:11:26	U	SMPL
J4699-04 - 1	CG-SOIL-2	1199	0.2619	-7471B	PPB	PPB	30 Aug 2018 17:13:43	U	SMPL
J4275-19LX5 - 1		423	0.0909	-7471B	PPB	PPB	30 Aug 2018 17:16:00	U	SMPL
J4275-19A - 1		19724	4.3449	-7471B	PPB	PPB	30 Aug 2018 17:18:18	U	SMPL
CCV46 - 1	CCV46	23013	5.0698	-7471B	PPB	PPB	30 Aug 2018 17:20:36	U	SMPL
CCB46 - 1	CCB46	-13	-0.0052	-7471B	PPB	PPB	30 Aug 2018 17:22:52	U	SMPL