



Analytical Summary Report

Analysis Method: 9040C
Parameter: Corrosivity
Run Number: LB97648

Analyst By : sweta

Supervisor Review By : AMANDEEP

Slope : 98.6

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2321
buffer solution pH 7 yellow	W2357
PH 10.01 BUFFER,COLOR CD 475ML	W2323
buffer solution pH 7 yellow	W2322
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			22.0	4.01	08/30/2018	16:30
2	CAL2	1	Water			21.8	7.01	08/30/2018	16:32
3	CAL3	1	Water			21.9	10.04	08/30/2018	16:34
4	ICV	1	Water			22.0	7.00	08/30/2018	16:36
5	CCV1	1	Water			21.8	2.00	08/30/2018	16:38
6	J4735-02	1	Water			28.3	7.32	08/30/2018	16:40
7	J4735-02DUP	1	Water			28.4	7.33	08/30/2018	16:42
8	J4736-01	1	Water			24.2	7.86	08/30/2018	16:44
9	CCV2	1	Water			21.9	12.00	08/30/2018	16:46

WORKLIST(Hardcopy Internal Chain)

LB 97648

WorkList Name : corrosivity083018-1

WorkList ID : 116245

Date : 8/30/2018 4:04:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
08/30/2018	Water	J4735-02	Corrosivity	Cool 4 deg C	PSEG01	E61	KG-BASELINE-WATER	08/30/2018	9040C
08/30/2018	Water	J4736-01	Corrosivity	Cool 4 deg C	AREC01	H13	WC-1	08/30/2018	9040C

Date/Time 08/30/18 4:15
 Received by: SP
 Relinquished by: CP

Date/Time 08/30/18 4:50
 Received by: CP
 Relinquished by: SP