


ISM02.4

LB 9765-9
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42087



A= 0.0000e+000

B= 2.4007e-004

C= -4.4429e-002

Rho= 0.9999348

Accept=Accepted

Slope
 $y = mx + b$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%SD
0.00	0.000	-0.013	-0.013	130	0.000	130					18
0.2	0.200	0.235	0.035	1162	0.0 %	1162					1
2.5	2.500	2.518	0.018	10674	0.0 %	10674					1
5.0	5.000	4.966	-0.034	20872	0.0 %	20872					1
7.5	7.500	7.435	-0.065	31153	0.0 %	31153					1
10.0	10.000	10.060	0.060	42087	0.0 %	42087					1

NB
8/31/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	130	-		0 ISM02.4	PPB	8/31/2018 8:13	Std
	0.2 S0.2	1162	-		0.2 ISM02.4	PPB	8/31/2018 8:15	Std
	2.5 S2.5	10674	-		2.5 ISM02.4	PPB	8/31/2018 8:17	Std
	5 S5	20872	-		5 ISM02.4	PPB	8/31/2018 8:20	Std
	7.5 S7.5	31153	-		7.5 ISM02.4	PPB	8/31/2018 8:22	Std
	10 S10.0	42087	-		10 ISM02.4	PPB	8/31/2018 8:24	Std
ICV	ICV	17039	4.0462 -		ISM02.4	PPB	8/31/2018 8:28	SMPL
ICB	ICB	15	-0.0408 -		ISM02.4	PPB	8/31/2018 8:30	SMPL
CCV047	CCV047	20991	4.995 -		ISM02.4	PPB	8/31/2018 8:32	SMPL
CCB047	CCB047	56	-0.031 -		ISM02.4	PPB	8/31/2018 8:35	SMPL
PB112511BL	PBW003	13	-0.0413 -		ISM02.4	PPB	8/31/2018 8:37	SMPL
J4714-01	MC0AE2	-26	-0.0507 -		ISM02.4	PPB	8/31/2018 8:39	SMPL
J4714-02	MC0AE3	-57	-0.0581 -		ISM02.4	PPB	8/31/2018 8:42	SMPL
J4714-03	MC0AE4	-50	-0.0564 -		ISM02.4	PPB	8/31/2018 8:44	SMPL
J4714-04	MC0AE5	-65	-0.06 -		ISM02.4	PPB	8/31/2018 8:46	SMPL
J4714-05	MC0AE6	-16	-0.0483 -		ISM02.4	PPB	8/31/2018 8:48	SMPL
J4714-06	MC0AE7	31	-0.037 -		ISM02.4	PPB	8/31/2018 8:51	SMPL
J4714-07	MC0AE7D	25	-0.0384 -		ISM02.4	PPB	8/31/2018 8:53	SMPL
J4714-08	MC0AE7S	4235	0.9723 -		ISM02.4	PPB	8/31/2018 8:55	SMPL
CCV048	CCV048	20909	4.9753 -		ISM02.4	PPB	8/31/2018 8:57	SMPL
CCB048	CCB048	29	-0.0375 -		ISM02.4	PPB	8/31/2018 9:00	SMPL