



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB97662

Ambient Barometric Pressure (mmHg): 760

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.0	80.6	80.6	08/31/2018	09:35
2	J4735-02		1	8	>100.0	>212.0	>212.0	08/31/2018	10:10
3	J4735-02DUP		1	8	>100.0	>212.0	>212.0	08/31/2018	10:40
4	J4736-01		1	9	>100.0	>212.0	>212.0	08/31/2018	11:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 97-662

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP83118

WorkList ID : 116272

Date : 8/31/2018 8:55:15 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
09/04/2018	Water	J4735-02 C	Flash Point	Cool 4 deg C	PSEG01	E61	KG-BASELINE-WATER	08/30/2018	1010A
09/13/2018	Water	J4736-01 K	Flash Point	Cool 4 deg C	AREC01	H13	WC-1	08/30/2018	1010A

Date/Time 08/31/18 9:00 AM
Received by: KS
Relinquished by: CP

Date/Time 08/31/18 11:20 AM
Received by: CP
Relinquished by: KS